

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY012423\
 Data File : VY012331.D
 Acq On : 24 Jan 2023 09:56
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 25 01:30:03 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011623S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 04:31:47 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	89	0.00
2 T	Dichlorodifluoromethane	0.437	0.380	13.0	91	0.00
3 P	Chloromethane	0.403	0.352	12.7	85	0.00
4 C	Vinyl Chloride	0.471	0.438	7.0#	90	0.00
5 T	Bromomethane	0.364	0.348	4.4	93	0.00
6 T	Chloroethane	0.307	0.301	2.0	93	0.00
7 T	Trichlorofluoromethane	0.919	0.910	1.0	96	0.00
8 T	Diethyl Ether	0.273	0.266	2.6	91	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.447	0.445	0.4	97	0.00
10 T	Methyl Iodide	0.680	0.619	9.0	81	0.00
11 T	Tert butyl alcohol	0.057	0.046	19.3	88	0.01
12 CM	1,1-Dichloroethene	0.459	0.452	1.5#	92	0.00
13 T	Acrolein	0.023	0.019	17.4	103	0.00
14 T	Allyl chloride	0.568	0.541	4.8	89	0.00
15 T	Acrylonitrile	0.119	0.121	-1.7	92	0.00
16 T	Acetone	0.139	0.153	-10.1	88	0.00
17 T	Carbon Disulfide	1.420	1.335	6.0	89	0.00
18 T	Methyl Acetate	0.312	0.302	3.2	89	0.00
19 T	Methyl tert-butyl Ether	1.280	1.296	-1.3	90	0.00
20 T	Methylene Chloride	0.762	0.500	34.4#	78	0.00
21 T	trans-1,2-Dichloroethene	0.512	0.513	-0.2	92	0.00
22 T	Diisopropyl ether	1.132	1.104	2.5	90	0.00
23 T	Vinyl Acetate	0.705	0.700	0.7	89	0.00
24 P	1,1-Dichloroethane	0.785	0.770	1.9	90	0.00
25 T	2-Butanone	0.154	0.158	-2.6	90	0.00
26 T	2,2-Dichloropropane	0.852	0.857	-0.6	94	0.00
27 T	cis-1,2-Dichloroethene	0.563	0.564	-0.2	92	0.00
28 T	Bromochloromethane	0.171	0.148	13.5	83	0.00
29 T	Tetrahydrofuran	0.084	0.085	-1.2	88	0.00
30 C	Chloroform	0.927	0.942	-1.6#	93	0.00
31 T	Cyclohexane	0.729	0.678	7.0	95	0.00
32 T	1,1,1-Trichloroethane	0.918	0.930	-1.3	94	0.00
33 S	1,2-Dichloroethane-d4	0.465	0.408	12.3	89	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	87	0.00
35 S	Dibromofluoromethane	0.265	0.242	8.7	88	0.00
36 T	1,1-Dichloropropene	0.450	0.454	-0.9	93	0.00
37 T	Ethyl Acetate	0.198	0.194	2.0	86	0.00
38 T	Carbon Tetrachloride	0.581	0.606	-4.3	96	0.00
39 T	Methylcyclohexane	0.550	0.574	-4.4	97	0.00
40 TM	Benzene	1.299	1.310	-0.8	92	0.00
41 T	Methacrylonitrile	0.117	0.121	-3.4	88	-0.01
42 TM	1,2-Dichloroethane	0.413	0.431	-4.4	92	0.00
43 T	Isopropyl Acetate	0.369	0.372	-0.8	88	0.00
44 TM	Trichloroethene	0.388	0.394	-1.5	92	0.00
45 C	1,2-Dichloropropane	0.271	0.273	-0.7#	91	0.00
46 T	Dibromomethane	0.181	0.187	-3.3	91	0.00
47 T	Bromodichloromethane	0.457	0.469	-2.6	92	0.00
48 T	Methyl methacrylate	0.170	0.173	-1.8	86	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY012423\
 Data File : VY012331.D
 Acq On : 24 Jan 2023 09:56
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 25 01:30:03 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011623S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 04:31:47 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.003	-50.0#	91	0.00
50 S	Toluene-d8	1.044	0.889	14.8	87	0.00
51 T	4-Methyl-2-Pentanone	0.197	0.205	-4.1	90	0.00
52 CM	Toluene	0.865	0.901	-4.2#	93	0.00
53 T	t-1,3-Dichloropropene	0.478	0.497	-4.0	91	0.00
54 T	cis-1,3-Dichloropropene	0.514	0.528	-2.7	91	0.00
55 T	1,1,2-Trichloroethane	0.258	0.261	-1.2	89	0.00
56 T	Ethyl methacrylate	0.338	0.354	-4.7	89	0.00
57 T	1,3-Dichloropropane	0.421	0.437	-3.8	92	0.00
58 T	2-Chloroethyl Vinyl ether	0.112	0.111	0.9	89	0.00
59 T	2-Hexanone	0.150	0.164	-9.3	90	0.00
60 T	Dibromochloromethane	0.342	0.356	-4.1	90	0.00
61 T	1,2-Dibromoethane	0.249	0.259	-4.0	91	0.00
62 S	4-Bromofluorobenzene	0.373	0.344	7.8	90	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	89	0.00
64 T	Tetrachloroethene	0.456	0.484	-6.1	95	0.00
65 PM	Chlorobenzene	1.023	1.037	-1.4	93	0.00
66 T	1,1,1,2-Tetrachloroethane	0.386	0.403	-4.4	93	0.00
67 C	Ethyl Benzene	1.828	1.881	-2.9#	93	0.00
68 T	m/p-Xylenes	0.723	0.750	-3.7	94	0.00
69 T	o-Xylene	0.690	0.708	-2.6	92	0.00
70 T	Styrene	1.154	1.205	-4.4	93	0.00
71 P	Bromoform	0.247	0.254	-2.8	91	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	89	0.00
73 T	Isopropylbenzene	3.605	3.743	-3.8	95	0.00
74 T	N-amyl acetate	0.707	0.688	2.7	86	0.00
75 P	1,1,2,2-Tetrachloroethane	0.568	0.577	-1.6	91	0.00
76 T	1,2,3-Trichloropropane	0.460	0.424	7.8	75	0.00
77 T	Bromobenzene	0.868	0.874	-0.7	91	0.00
78 T	n-propylbenzene	4.219	4.301	-1.9	95	0.00
79 T	2-Chlorotoluene	2.385	2.435	-2.1	94	0.00
80 T	1,3,5-Trimethylbenzene	3.134	3.231	-3.1	95	0.00
81 T	trans-1,4-Dichloro-2-butene	0.223	0.228	-2.2	90	0.00
82 T	4-Chlorotoluene	2.506	2.509	-0.1	93	0.00
83 T	tert-Butylbenzene	2.732	2.858	-4.6	99	0.00
84 T	1,2,4-Trimethylbenzene	3.091	3.207	-3.8	94	0.00
85 T	sec-Butylbenzene	3.889	4.016	-3.3	97	0.00
86 T	p-Isopropyltoluene	3.398	3.538	-4.1	97	0.00
87 T	1,3-Dichlorobenzene	1.725	1.726	-0.1	92	0.00
88 T	1,4-Dichlorobenzene	1.721	1.706	0.9	93	0.00
89 T	n-Butylbenzene	2.945	3.042	-3.3	97	0.00
90 T	Hexachloroethane	0.613	0.609	0.7	93	0.00
91 T	1,2-Dichlorobenzene	1.540	1.540	0.0	92	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.117	0.122	-4.3	93	0.00
93 T	1,2,4-Trichlorobenzene	0.987	0.975	1.2	93	0.00
94 T	Hexachlorobutadiene	0.594	0.598	-0.7	97	0.00
95 T	Naphthalene	1.947	1.993	-2.4	92	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY012423\
Data File : VY012331.D
Acq On : 24 Jan 2023 09:56
Operator : KP/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Jan 25 01:30:03 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011623S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 17 04:31:47 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.849	0.838	1.3	94	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6