



Data Path : Z:\voasrv\HPCHEM1\MSVOA\_Y\Data\VY012424\  
Data File : VY017084.D  
Acq On : 24 Jan 2024 09:35  
Operator : SY/MD  
Sample : BFB  
Misc : 5.00g/5.0mL/MSVOA\_Y/SOIL  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
MSVOA\_Y  
ClientSampleId :  
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA\_Y\methods\82Y012424S.M  
Title : SW846 8260  
Last Update : Thu Jan 25 04:44:10 2024



AutoFind: Scans 1750, 1751, 1752; Background Corrected with Scan 1742

| Target<br>Mass | Rel. to<br>Mass | Lower<br>Limit% | Upper<br>Limit% | Rel.<br>Abn% | Raw<br>Abn | Result<br>Pass/Fail |
|----------------|-----------------|-----------------|-----------------|--------------|------------|---------------------|
| 50             | 95              | 15              | 40              | 21.0         | 13376      | PASS                |
| 75             | 95              | 30              | 60              | 53.6         | 34051      | PASS                |
| 95             | 95              | 100             | 100             | 100.0        | 63560      | PASS                |
| 96             | 95              | 5               | 9               | 6.9          | 4414       | PASS                |
| 173            | 174             | 0.00            | 2               | 1.3          | 633        | PASS                |
| 174            | 95              | 50              | 100             | 78.0         | 49608      | PASS                |
| 175            | 174             | 5               | 9               | 7.7          | 3830       | PASS                |
| 176            | 174             | 95              | 101             | 96.7         | 47971      | PASS                |
| 177            | 176             | 5               | 9               | 6.5          | 3132       | PASS                |