

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY012523\
 Data File : VY012362.D
 Acq On : 25 Jan 2023 17:28
 Operator : KP/MD
 Sample : 01243-05RE
 Misc : 6.19g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TJRC-011923-B4-20-30RE

Quant Time: Jan 26 04:10:11 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011623S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 04:31:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	3893	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	6188	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	5820	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	1584	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	2896	93.139	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 186.280%#	
35) Dibromofluoromethane	7.710	113	974	29.726	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 59.460%	
50) Toluene-d8	10.179	98	7187	64.569	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 129.140%	
62) 4-Bromofluorobenzene	12.483	95	1970	42.651	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 85.300%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY012523\
Data File : VY012362.D
Acq On : 25 Jan 2023 17:28
Operator : KP/MD
Sample : 01243-05RE
Misc : 6.19g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TJRC-011923-B4-20-30RE

Quant Time: Jan 26 04:10:11 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011623S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 17 04:31:47 2023
Response via : Initial Calibration

