

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY012624\
 Data File : VY017113.D
 Acq On : 26 Jan 2024 11:07
 Operator : SY/MD
 Sample : VY0126SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0126SBL01

Quant Time: Jan 29 04:14:19 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y012424S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 25 04:44:10 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	99647	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	183868	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	157644	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	61936	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	63862	65.778	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	131.560%
35) Dibromofluoromethane	7.722	113	64580	58.195	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	116.380%
50) Toluene-d8	10.185	98	233008	55.591	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	111.180%
62) 4-Bromofluorobenzene	12.483	95	67973	52.643	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	105.280%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY012624\
Data File : VY017113.D
Acq On : 26 Jan 2024 11:07
Operator : SY/MD
Sample : VY0126SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0126SBL01

Quant Time: Jan 29 04:14:19 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y012424S.M
Quant Title : SW846 8260
QLast Update : Thu Jan 25 04:44:10 2024
Response via : Initial Calibration

