



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY012725\
Data File : VY020935.D
Acq On : 27 Jan 2025 08:32
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y012425S.M
Title : SW846 8260
Last Update : Sat Jan 25 01:12:18 2025



AutoFind: Scans 1754, 1755, 1756; Background Corrected with Scan 1746

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.4	19955	PASS
75	95	30	60	49.7	60339	PASS
95	95	100	100	100.0	121421	PASS
96	95	5	9	6.8	8277	PASS
173	174	0.00	2	1.1	1129	PASS
174	95	50	100	86.4	104901	PASS
175	174	5	9	7.7	8049	PASS
176	174	95	101	96.7	101461	PASS
177	176	5	9	7.2	7294	PASS