

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY012822\
 Data File : VY007319.D
 Acq On : 28 Jan 2022 14:35
 Operator : SY/MD
 Sample : N1309-01
 Misc : 5.69g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 STATE-ST-SP1-(1)

Quant Time: Jan 31 00:30:23 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011322S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jan 14 05:48:29 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	106888	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	199499	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	189386	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	82067	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	65829	50.193	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	100.380%
35) Dibromofluoromethane	7.722	113	66586	49.360	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	98.720%
50) Toluene-d8	10.179	98	244815	49.667	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	99.340%
62) 4-Bromofluorobenzene	12.483	95	87423	49.860	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	99.720%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY012822\
Data File : VY007319.D
Acq On : 28 Jan 2022 14:35
Operator : SY/MD
Sample : N1309-01
Misc : 5.69g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
STATE-ST-SP1-(1)

Quant Time: Jan 31 00:30:23 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011322S.M
Quant Title : SW846 8260
QLast Update : Fri Jan 14 05:48:29 2022
Response via : Initial Calibration

