

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY012822\
 Data File : VY007325.D
 Acq On : 28 Jan 2022 16:55
 Operator : SY/MD
 Sample : N1309-07
 Misc : 5.26g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 STATE-ST-SP1-(7)

Quant Time: Jan 31 00:31:12 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011322S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jan 14 05:48:29 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	104054	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	194508	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	184194	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	79451	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	65479	51.285	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 102.580%	
35) Dibromofluoromethane	7.716	113	64947	49.381	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 98.760%	
50) Toluene-d8	10.179	98	239619	49.860	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 99.720%	
62) 4-Bromofluorobenzene	12.483	95	88814	51.953	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 103.900%	
Target Compounds						Qvalue
16) Acetone	3.948	43	6292	18.989	ug/l	100
20) Methylene Chloride	4.716	84	15575	11.367	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY012822\
Data File : VY007325.D
Acq On : 28 Jan 2022 16:55
Operator : SY/MD
Sample : N1309-07
Misc : 5.26g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
STATE-ST-SP1-(7)

Quant Time: Jan 31 00:31:12 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011322S.M
Quant Title : SW846 8260
QLast Update : Fri Jan 14 05:48:29 2022
Response via : Initial Calibration

