

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY012822\
 Data File : VY007328.D
 Acq On : 28 Jan 2022 18:06
 Operator : SY/MD
 Sample : N1309-10
 Misc : 4.67g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 STATE-ST-SP1-(10)

Quant Time: Jan 31 00:31:37 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011322S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jan 14 05:48:29 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	103770	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	190474	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	182450	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	80154	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	64825	50.912	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	101.820%
35) Dibromofluoromethane	7.722	113	63466	49.276	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	98.560%
50) Toluene-d8	10.179	98	235619	50.066	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	100.140%
62) 4-Bromofluorobenzene	12.483	95	88029	52.585	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	105.160%
Target Compounds						
18) Methyl Acetate	4.466	43	2759	2.346	ug/l #	67
20) Methylene Chloride	4.704	84	10859	7.947	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY012822\
Data File : VY007328.D
Acq On : 28 Jan 2022 18:06
Operator : SY/MD
Sample : N1309-10
Misc : 4.67g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
STATE-ST-SP1-(10)

Quant Time: Jan 31 00:31:37 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011322S.M
Quant Title : SW846 8260
QLast Update : Fri Jan 14 05:48:29 2022
Response via : Initial Calibration

