


Data Path : Z:\voasrv\HPCHEM1\MSVOA Y\Data\vv012920\
 Data File : VY001412.D
 Acq On : 29 Jan 2020 10:16
 Operator : SY/MD
 Sample : BFB
 Misc : 5.00G/5ML/MSVOA Y/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_Y\METHODS\82Y012120S.M
 Title : SW846 8260
 Last Update : Thu Jan 23 01:21:47 2020



AutoFind: Scans 1764, 1765, 1766; Background Corrected with Scan 1756

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.3	18402	PASS
75	95	30	60	49.5	49810	PASS
95	95	100	100	100.0	100688	PASS
96	95	5	9	6.7	6712	PASS
173	174	0.00	2	0.8	589	PASS
174	95	50	100	74.1	74592	PASS
175	174	5	9	7.2	5386	PASS
176	174	95	101	96.3	71824	PASS
177	176	5	9	7.3	5235	PASS