

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY012925\
 Data File : VY020958.D
 Acq On : 29 Jan 2025 09:30
 Operator : SY/MD
 Sample : VY0129SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0129SBL01

Quant Time: Jan 30 00:31:43 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y012425S.M
 Quant Title : SW846 8260
 QLast Update : Sat Jan 25 01:12:18 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.719	168	210912	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.622	114	361800	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.426	117	309527	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.359	152	119372	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.073	65	105932	62.994	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	125.980%
35) Dibromofluoromethane	7.646	113	114754	55.249	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	110.500%
50) Toluene-d8	10.115	98	420548	52.200	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	104.400%
62) 4-Bromofluorobenzene	12.414	95	126487	47.175	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	94.340%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY012925\
Data File : VY020958.D
Acq On : 29 Jan 2025 09:30
Operator : SY/MD
Sample : VY0129SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0129SBL01

Quant Time: Jan 30 00:31:43 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y012425S.M
Quant Title : SW846 8260
QLast Update : Sat Jan 25 01:12:18 2025
Response via : Initial Calibration

