

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY012925\
 Data File : VY020962.D
 Acq On : 29 Jan 2025 11:48
 Operator : SY/MD
 Sample : Q1205-01
 Misc : 4.59g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VNJ-236

Quant Time: Jan 30 00:34:24 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y012425S.M
 Quant Title : SW846 8260
 QLast Update : Sat Jan 25 01:12:18 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	372623	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.615	114	571823	50.000	ug/l	-0.01
63) Chlorobenzene-d5	11.420	117	457613	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.352	152	198870	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.067	65	145874	49.100	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	98.200%
35) Dibromofluoromethane	7.634	113	161809	49.291	ug/l	-0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	98.580%
50) Toluene-d8	10.109	98	608406	47.781	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	95.560%
62) 4-Bromofluorobenzene	12.408	95	182785	43.133	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	86.260%

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY012925\
Data File : VY020962.D
Acq On : 29 Jan 2025 11:48
Operator : SY/MD
Sample : Q1205-01
Misc : 4.59g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VNJ-236

Quant Time: Jan 30 00:34:24 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y012425S.M
Quant Title : SW846 8260
QLast Update : Sat Jan 25 01:12:18 2025
Response via : Initial Calibration

