

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY013024\
 Data File : VY017133.D
 Acq On : 30 Jan 2024 09:31
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :

Quant Time: Jan 31 00:26:08 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y012424S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 25 04:44:10 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.813	168	118450	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	8.710	114	195320	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.508	117	166967	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.440	152	72856	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.167	65	61104	52.947	ug/l	0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	= 105.900%	
35) Dibromofluoromethane	7.740	113	66819	56.682	ug/l	0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	= 113.360%	
50) Toluene-d8	10.197	98	253442	56.921	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 113.840%	
62) 4-Bromofluorobenzene	12.502	95	76468	55.750	ug/l	0.01
Spiked Amount	50.000	Range	30 - 143	Recovery	= 111.500%	
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.918	85	40648	47.037	ug/l	97
3) Chloromethane	2.131	50	72863	48.492	ug/l	100
4) Vinyl Chloride	2.266	62	87792	51.569	ug/l	100
5) Bromomethane	2.674	94	62841	49.237	ug/l	97
6) Chloroethane	2.820	64	60018	53.696	ug/l	99
7) Trichlorofluoromethane	3.150	101	100962	51.952	ug/l	99
8) Diethyl Ether	3.552	74	30699	44.982	ug/l	95
9) 1,1,2-Trichlorotrifluo...	3.930	101	63889	48.222	ug/l	100
10) Methyl Iodide	4.119	142	62067	49.311	ug/l	97
11) Tert butyl alcohol	4.991	59	17444	202.829	ug/l	98
12) 1,1-Dichloroethene	3.899	96	58527	48.258	ug/l	97
13) Acrolein	3.753	56	27176	196.347	ug/l	98
14) Allyl chloride	4.509	41	90936	49.607	ug/l	99
15) Acrylonitrile	5.198	53	68912	232.659	ug/l	100
16) Acetone	3.973	43	71833	227.853	ug/l	97
17) Carbon Disulfide	4.229	76	160751	43.977	ug/l	99
18) Methyl Acetate	4.509	43	60062	50.732	ug/l	100
19) Methyl tert-butyl Ether	5.253	73	137712	46.532	ug/l	99
20) Methylene Chloride	4.747	84	66977	52.494	ug/l	96
21) trans-1,2-Dichloroethene	5.259	96	67336	48.001	ug/l	98
22) Diisopropyl ether	6.149	45	207880	51.375	ug/l	98
23) Vinyl Acetate	6.094	43	458686	233.690	ug/l	100
24) 1,1-Dichloroethane	6.051	63	124166	50.309	ug/l	99
25) 2-Butanone	7.015	43	93086	214.024	ug/l	99
26) 2,2-Dichloropropane	7.009	77	104351	49.503	ug/l	100
27) cis-1,2-Dichloroethene	7.015	96	78555	50.533	ug/l	100
28) Bromochloromethane	7.362	49	53070	54.012	ug/l	99
29) Tetrahydrofuran	7.374	42	53682	229.394	ug/l	98
30) Chloroform	7.533	83	125972	50.126	ug/l	100
31) Cyclohexane	7.813	56	102906	50.278	ug/l	95
32) 1,1,1-Trichloroethane	7.728	97	108451	50.326	ug/l	100
36) 1,1-Dichloropropene	7.941	75	97309	50.170	ug/l	99
37) Ethyl Acetate	7.100	43	37933	45.484	ug/l	99
38) Carbon Tetrachloride	7.929	117	95623	50.309	ug/l	100
39) Methylcyclohexane	9.203	83	114474	49.071	ug/l	98
40) Benzene	8.185	78	285487	50.683	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY013024\
 Data File : VY017133.D
 Acq On : 30 Jan 2024 09:31
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :

Quant Time: Jan 31 00:26:08 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y012424S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 25 04:44:10 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.338	41	22257	49.900	ug/l	98
42) 1,2-Dichloroethane	8.258	62	70896	48.548	ug/l	100
43) Isopropyl Acetate	8.295	43	72489	46.559	ug/l	99
44) Trichloroethene	8.960	130	72148	49.940	ug/l	97
45) 1,2-Dichloropropane	9.234	63	71195	51.496	ug/l	99
46) Dibromomethane	9.325	93	35316	48.099	ug/l	99
47) Bromodichloromethane	9.514	83	94755	50.727	ug/l	97
48) Methyl methacrylate	9.313	41	40300	46.619	ug/l	99
49) 1,4-Dioxane	9.319	88	7288	1009.236	ug/l	96
51) 4-Methyl-2-Pentanone	10.087	43	195133	235.609	ug/l	99
52) Toluene	10.264	92	175298	51.417	ug/l	100
53) t-1,3-Dichloropropene	10.484	75	84929	49.771	ug/l	98
54) cis-1,3-Dichloropropene	9.947	75	105923	50.470	ug/l	98
55) 1,1,2-Trichloroethane	10.660	97	49757	48.487	ug/l	99
56) Ethyl methacrylate	10.526	69	42733	47.619	ug/l	97
57) 1,3-Dichloropropane	10.807	76	85954	49.542	ug/l	99
58) 2-Chloroethyl Vinyl ether	9.801	63	184161	295.642	ug/l	100
59) 2-Hexanone	10.849	43	136750	216.429	ug/l	99
60) Dibromochloromethane	11.002	129	58401	48.985	ug/l	100
61) 1,2-Dibromoethane	11.105	107	44600	47.842	ug/l	100
64) Tetrachloroethene	10.740	164	56944	48.929	ug/l	95
65) Chlorobenzene	11.532	112	177253	50.096	ug/l	99
66) 1,1,1,2-Tetrachloroethane	11.605	131	66471	51.432	ug/l	99
67) Ethyl Benzene	11.611	91	334257	51.966	ug/l	97
68) m/p-Xylenes	11.721	106	254690	105.388	ug/l	98
69) o-Xylene	12.044	106	113772	51.827	ug/l	100
70) Styrene	12.063	104	170306	53.439	ug/l	99
71) Bromoform	12.227	173	32080	49.502	ug/l #	98
73) Isopropylbenzene	12.349	105	318437	53.167	ug/l	100
74) N-amyl acetate	12.160	43	56673	46.944	ug/l	99
75) 1,1,2,2-Tetrachloroethane	12.599	83	54399	47.631	ug/l	99
76) 1,2,3-Trichloropropane	12.648	75	66396	75.081	ug/l #	100
77) Bromobenzene	12.623	156	66487	51.303	ug/l	97
78) n-propylbenzene	12.691	91	384134	53.324	ug/l	100
79) 2-Chlorotoluene	12.776	91	207350	52.452	ug/l	98
80) 1,3,5-Trimethylbenzene	12.831	105	251005	52.953	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.398	75	16066	47.034	ug/l	98
82) 4-Chlorotoluene	12.873	91	213330	52.069	ug/l	100
83) tert-Butylbenzene	13.093	119	220389	53.752	ug/l	99
84) 1,2,4-Trimethylbenzene	13.136	105	247437	52.520	ug/l	98
85) sec-Butylbenzene	13.270	105	325159	53.559	ug/l	99
86) p-Isopropyltoluene	13.386	119	271394	53.666	ug/l	100
87) 1,3-Dichlorobenzene	13.386	146	131279	51.307	ug/l	99
88) 1,4-Dichlorobenzene	13.465	146	124824	49.583	ug/l	99
89) n-Butylbenzene	13.715	91	260743	53.454	ug/l	99
90) Hexachloroethane	13.977	117	54724	51.918	ug/l	100
91) 1,2-Dichlorobenzene	13.757	146	109975	50.330	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.373	75	6401	43.883	ug/l	99
93) 1,2,4-Trichlorobenzene	15.025	180	55782	48.852	ug/l	98
94) Hexachlorobutadiene	15.129	225	33609	48.906	ug/l	99
95) Naphthalene	15.257	128	94815	47.014	ug/l	100
96) 1,2,3-Trichlorobenzene	15.446	180	45034	47.405	ug/l	97


Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY013024\
Data File : VY017133.D
Acq On : 30 Jan 2024 09:31
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :

Quant Time: Jan 31 00:26:08 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y012424S.M
Quant Title : SW846 8260
QLast Update : Thu Jan 25 04:44:10 2024
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed






#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.813 min Scan# 9
Delta R.T. 0.012 min
Lab File: VY017133.D
Acq: 30 Jan 2024 09:31

Instrument :
MSVOA_Y
ClientSampleId :






Tgt Ion:168 Resp: 118450
Ion Ratio Lower Upper
168 100
99 53.2 44.2 66.4






#2
Dichlorodifluoromethane
Concen: 47.037 ug/l
RT: 1.918 min Scan# 16
Delta R.T. 0.006 min
Lab File: VY017133.D
Acq: 30 Jan 2024 09:31

Tgt Ion: 85 Resp: 40648
Ion Ratio Lower Upper
85 100
87 32.4 17.1 51.3







Abundance Scan 49 (2.119 min): VY017088.D\data.ms (-43)



#3
Chloromethane
Concen: 48.492 ug/l
RT: 2.131 min Scan# 51
Delta R.T. 0.012 min
Lab File: VY017133.D
Acq: 30 Jan 2024 09:31


Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion: 50 Resp: 72863
Ion Ratio Lower Upper
50 100
52 33.5 26.7 40.1



#4
Vinyl Chloride
Concen: 51.569 ug/l
RT: 2.266 min Scan# 73
Delta R.T. 0.007 min
Lab File: VY017133.D
Acq: 30 Jan 2024 09:31

Tgt Ion: 62 Resp: 87792
Ion Ratio Lower Upper
62 100
64 32.3 25.7 38.5









#5
Bromomethane
Concen: 49.237 ug/l
RT: 2.674 min Scan# 14
Delta R.T. 0.012 min
Lab File: VY017133.D
Acq: 30 Jan 2024 09:31




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion: 94 Resp: 62841
Ion Ratio Lower Upper
94 100
96 93.6 77.3 115.9



#6
Chloroethane
Concen: 53.696 ug/l
RT: 2.820 min Scan# 164
Delta R.T. 0.012 min
Lab File: VY017133.D
Acq: 30 Jan 2024 09:31

Tgt Ion: 64 Resp: 60018
Ion Ratio Lower Upper
64 100
66 32.1 25.1 37.7









#7
Trichlorofluoromethane
Concen: 51.952 ug/l
RT: 3.150 min Scan# 216
Delta R.T. 0.013 min
Lab File: VY017133.D
Acq: 30 Jan 2024 09:31




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion:101 Resp: 100962
Ion Ratio Lower Upper
101 100
103 63.9 51.9 77.9



#8
Diethyl Ether
Concen: 44.982 ug/l
RT: 3.552 min Scan# 284
Delta R.T. 0.012 min
Lab File: VY017133.D
Acq: 30 Jan 2024 09:31

Tgt Ion: 74 Resp: 30699
Ion Ratio Lower Upper
74 100
45 100.8 47.9 143.7









#9
1,1,2-Trichlorotrifluoroethane
Concen: 48.222 ug/l
RT: 3.930 min Scan# 344
Delta R.T. 0.019 min
Lab File: VY017133.D
Acq: 30 Jan 2024 09:31




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion:101 Resp: 63889
Ion Ratio Lower Upper
101 100
85 45.0 35.8 53.6
151 72.9 58.6 88.0



#10
Methyl Iodide
Concen: 49.311 ug/l
RT: 4.119 min Scan# 377
Delta R.T. 0.013 min
Lab File: VY017133.D
Acq: 30 Jan 2024 09:31

Tgt Ion:142 Resp: 62067
Ion Ratio Lower Upper
142 100
127 41.8 35.1 52.7
141 14.6 11.8 17.6





#11

Tert butyl alcohol

Concen: 202.829 ug/l

RT: 4.991 min Scan# 51

Delta R.T. 0.025 min

Lab File: VY017133.D

Acq: 30 Jan 2024 09:31

Instrument :

MSVOA_Y





ClientSampleId :

Tgt Ion: 59 Resp: 17444

Ion Ratio Lower Upper

59 100

57 9.7 7.4 11.0



#12

1,1-Dichloroethene

Concen: 48.258 ug/l

RT: 3.899 min Scan# 341

Delta R.T. 0.012 min

Lab File: VY017133.D

Acq: 30 Jan 2024 09:31




Tgt Ion: 96 Resp: 58527


Ion Ratio Lower Upper

96 100

61 158.8 124.2 186.2


98 63.5 52.6 79.0








#13
Acrolein
Concen: 196.347 ug/l
RT: 3.753 min Scan# 315
Delta R.T. 0.012 min
Lab File: VY017133.D
Acq: 30 Jan 2024 09:31


Instrument :
MSVOA_Y
ClientSampleId :





Tgt Ion: 56 Resp: 27176
Ion Ratio Lower Upper
56 100
55 70.6 55.4 83.0




#14
Allyl chloride
Concen: 49.607 ug/l
RT: 4.509 min Scan# 441
Delta R.T. 0.012 min
Lab File: VY017133.D
Acq: 30 Jan 2024 09:31




Tgt Ion: 41 Resp: 90936
Ion Ratio Lower Upper
41 100
39 69.6 54.7 82.1
76 39.4 31.8 47.6








#15
Acrylonitrile
Concen: 232.659 ug/l
RT: 5.198 min Scan# 51
Delta R.T. 0.019 min
Lab File: VY017133.D
Acq: 30 Jan 2024 09:31

Instrument :
MSVOA_Y
ClientSampleId :







Tgt Ion: 53 Resp: 68912
Ion Ratio Lower Upper
53 100
52 81.2 65.1 97.7
51 36.6 29.0 43.4



#16
Acetone
Concen: 227.853 ug/l
RT: 3.973 min Scan# 353
Delta R.T. 0.013 min
Lab File: VY017133.D
Acq: 30 Jan 2024 09:31

Tgt Ion: 43 Resp: 71833
Ion Ratio Lower Upper
43 100
58 31.4 26.6 39.8









#17
Carbon Disulfide
Concen: 43.977 ug/l
RT: 4.229 min Scan# 393
Delta R.T. 0.019 min
Lab File: VY017133.D
Acq: 30 Jan 2024 09:31




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion: 76 Resp: 160751
Ion Ratio Lower Upper
76 100
78 8.8 7.4 11.0



#18
Methyl Acetate
Concen: 50.732 ug/l
RT: 4.509 min Scan# 441
Delta R.T. 0.012 min
Lab File: VY017133.D
Acq: 30 Jan 2024 09:31

Tgt Ion: 43 Resp: 60062
Ion Ratio Lower Upper
43 100
74 24.5 19.8 29.6





#19

Methyl tert-butyl Ether

Concen: 46.532 ug/l

RT: 5.253 min Scan# 5

Delta R.T. 0.019 min

Lab File: VY017133.D

Acq: 30 Jan 2024 09:31

Instrument :

MSVOA_Y





ClientSampleId :

Tgt Ion: 73 Resp: 137712

Ion Ratio Lower Upper

73 100

57 22.7 17.7 26.5



#20

Methylene Chloride

Concen: 52.494 ug/l

RT: 4.747 min Scan# 480

Delta R.T. 0.013 min

Lab File: VY017133.D

Acq: 30 Jan 2024 09:31

Tgt Ion: 84 Resp: 66977




Ion Ratio Lower Upper


84 100

49 120.8 95.2 142.8

51 35.4 29.9 44.9


86 58.4 52.2 78.4








#21
trans-1,2-Dichloroethene
Concen: 48.001 ug/l
RT: 5.259 min Scan# 51
Delta R.T. 0.019 min
Lab File: VY017133.D
Acq: 30 Jan 2024 09:31


Instrument :
MSVOA_Y
ClientSampleId :





Tgt Ion: 96 Resp: 67336
Ion Ratio Lower Upper
96 100
61 128.5 104.6 156.8
98 62.1 50.9 76.3




#22
Diisopropyl ether
Concen: 51.375 ug/l
RT: 6.149 min Scan# 710
Delta R.T. 0.018 min
Lab File: VY017133.D
Acq: 30 Jan 2024 09:31



Tgt Ion: 45 Resp: 207880
Ion Ratio Lower Upper
45 100
43 55.6 46.8 70.2
87 27.6 22.2 33.2
59 12.1 9.9 14.9









#23
 Vinyl Acetate
 Concen: 233.690 ug/l
 RT: 6.094 min Scan# 701
 Delta R.T. 0.018 min
 Lab File: VY017133.D
 Acq: 30 Jan 2024 09:31




Instrument :
 MSVOA_Y
 ClientSampleId :


Tgt Ion: 43 Resp: 458686
 Ion Ratio Lower Upper
 43 100
 86 10.8 8.6 13.0



#24
 1,1-Dichloroethane
 Concen: 50.309 ug/l
 RT: 6.051 min Scan# 694
 Delta R.T. 0.018 min
 Lab File: VY017133.D
 Acq: 30 Jan 2024 09:31

Tgt Ion: 63 Resp: 124166
 Ion Ratio Lower Upper
 63 100
 98 6.9 3.8 11.4
 100 4.3 2.1 6.2





#25

2-Butanone

Concen: 214.024 ug/l

RT: 7.015 min Scan# 81

Delta R.T. 0.019 min

Lab File: VY017133.D

Acq: 30 Jan 2024 09:31

Instrument :

MSVOA_Y





ClientSampleId :

Tgt Ion: 43 Resp: 93086

Ion Ratio Lower Upper

43 100

72 26.0 21.4 32.2



#26

2,2-Dichloropropane

Concen: 49.503 ug/l

RT: 7.009 min Scan# 851

Delta R.T. 0.013 min

Lab File: VY017133.D




Acq: 30 Jan 2024 09:31


Tgt Ion: 77 Resp: 104351

Ion Ratio Lower Upper

77 100


97 23.5 11.7 35.0








#27
 cis-1,2-Dichloroethene
 Concen: 50.533 ug/l
 RT: 7.015 min Scan# 850
 Delta R.T. 0.013 min
 Lab File: VY017133.D
 Acq: 30 Jan 2024 09:31

Instrument :
 MSVOA_Y
 ClientSampleId :



Tgt Ion: 96 Resp: 78555




Ion	Ratio	Lower	Upper
96	100		
61	139.1	0.0	278.8
98	63.5	0.0	127.0




#28
 Bromochloromethane
 Concen: 54.012 ug/l
 RT: 7.362 min Scan# 909
 Delta R.T. 0.012 min
 Lab File: VY017133.D
 Acq: 30 Jan 2024 09:31

Tgt Ion: 49 Resp: 53070

Ion	Ratio	Lower	Upper
49	100		
129	2.1	0.0	3.8
130	73.9	59.8	89.8









#29
Tetrahydrofuran
Concen: 229.394 ug/l
RT: 7.374 min Scan# 911
Delta R.T. 0.012 min
Lab File: VY017133.D
Acq: 30 Jan 2024 09:31




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion: 42 Resp: 53682
Ion Ratio Lower Upper
42 100
72 45.4 36.6 54.8
71 40.3 33.8 50.6



#30
Chloroform
Concen: 50.126 ug/l
RT: 7.533 min Scan# 937
Delta R.T. 0.012 min
Lab File: VY017133.D
Acq: 30 Jan 2024 09:31


Tgt Ion: 83 Resp: 125972
Ion Ratio Lower Upper
83 100
85 63.6 51.2 76.8








#31
Cyclohexane
Concen: 50.278 ug/l
RT: 7.813 min Scan# 91
Delta R.T. 0.012 min
Lab File: VY017133.D
Acq: 30 Jan 2024 09:31

Instrument :
MSVOA_Y
ClientSampleId :







Tgt Ion: 56 Resp: 102906
Ion Ratio Lower Upper
56 100
69 32.9 27.0 40.6
84 95.0 71.8 107.6



#32
1,1,1-Trichloroethane
Concen: 50.326 ug/l
RT: 7.728 min Scan# 969
Delta R.T. 0.012 min
Lab File: VY017133.D
Acq: 30 Jan 2024 09:31

Tgt Ion: 97 Resp: 108451
Ion Ratio Lower Upper
97 100
99 64.0 51.2 76.8
61 44.5 35.0 52.6









#33
1,2-Dichloroethane-d4
Concen: 52.947 ug/l
RT: 8.167 min Scan# 1104
Delta R.T. 0.012 min
Lab File: VY017133.D
Acq: 30 Jan 2024 09:31




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion: 65 Resp: 61104
Ion Ratio Lower Upper
65 100
67 53.7 0.0 109.0



#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 8.710 min Scan# 1130
Delta R.T. 0.007 min
Lab File: VY017133.D
Acq: 30 Jan 2024 09:31

Tgt Ion: 114 Resp: 195320
Ion Ratio Lower Upper
114 100
63 19.3 0.0 37.4
88 14.5 0.0 29.0









#35
Dibromofluoromethane
Concen: 56.682 ug/l
RT: 7.740 min Scan# 91
Delta R.T. 0.012 min
Lab File: VY017133.D
Acq: 30 Jan 2024 09:31




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion:113 Resp: 66819
Ion Ratio Lower Upper
113 100
111 102.6 83.8 125.6
192 18.1 14.5 21.7



#36
1,1-Dichloropropene
Concen: 50.170 ug/l
RT: 7.941 min Scan# 1004
Delta R.T. 0.012 min
Lab File: VY017133.D
Acq: 30 Jan 2024 09:31


Tgt Ion: 75 Resp: 97309
Ion Ratio Lower Upper
75 100
110 34.3 17.4 52.3
77 31.3 24.7 37.1








#37
Ethyl Acetate
Concen: 45.484 ug/l
RT: 7.100 min Scan# 864
Delta R.T. 0.012 min
Lab File: VY017133.D
Acq: 30 Jan 2024 09:31

Instrument :
MSVOA_Y
ClientSampleId :







Tgt Ion: 43 Resp: 37933
Ion Ratio Lower Upper
43 100
61 13.5 11.0 16.6
70 9.4 8.1 12.1



#38
Carbon Tetrachloride
Concen: 50.309 ug/l
RT: 7.929 min Scan# 1002
Delta R.T. 0.012 min
Lab File: VY017133.D
Acq: 30 Jan 2024 09:31

Tgt Ion: 117 Resp: 95623
Ion Ratio Lower Upper
117 100
119 94.4 76.0 114.0
121 30.1 24.2 36.2





#39

Methylcyclohexane

Concen: 49.071 ug/l

RT: 9.203 min Scan# 11

Delta R.T. 0.012 min

Lab File: VY017133.D

Acq: 30 Jan 2024 09:31

Instrument :

MSVOA_Y

ClientSampleId :





Tgt Ion: 83 Resp: 114474

Ion Ratio Lower Upper

83 100

55 71.4 59.7 89.5

98 47.2 38.0 57.0



#40

Benzene

Concen: 50.683 ug/l

RT: 8.185 min Scan# 1044

Delta R.T. 0.012 min

Lab File: VY017133.D




Acq: 30 Jan 2024 09:31


Tgt Ion: 78 Resp: 285487

Ion Ratio Lower Upper

78 100


77 23.3 18.8 28.2








#41
Methacrylonitrile
Concen: 49.900 ug/l
RT: 7.338 min Scan# 903
Delta R.T. 0.019 min
Lab File: VY017133.D
Acq: 30 Jan 2024 09:31

Instrument :
MSVOA_Y
ClientSampleId :







Tgt Ion: 41 Resp: 22257
Ion Ratio Lower Upper
41 100
39 53.6 41.9 62.9
67 75.9 59.8 89.6
52 29.9 24.8 37.2



#42
1,2-Dichloroethane
Concen: 48.548 ug/l
RT: 8.258 min Scan# 1056
Delta R.T. 0.006 min
Lab File: VY017133.D
Acq: 30 Jan 2024 09:31

Tgt Ion: 62 Resp: 70896
Ion Ratio Lower Upper
62 100
98 10.3 0.0 20.2









#43
Isopropyl Acetate
Concen: 46.559 ug/l
RT: 8.295 min Scan# 1062
Delta R.T. 0.012 min
Lab File: VY017133.D
Acq: 30 Jan 2024 09:31




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion: 43 Resp: 72489
Ion Ratio Lower Upper
43 100
61 31.3 24.8 37.2
87 12.7 10.7 16.1



#44
Trichloroethene
Concen: 49.940 ug/l
RT: 8.960 min Scan# 1171
Delta R.T. 0.013 min
Lab File: VY017133.D
Acq: 30 Jan 2024 09:31

Tgt Ion:130 Resp: 72148
Ion Ratio Lower Upper
130 100
95 98.6 0.0 203.8





#45

1,2-Dichloropropane

Concen: 51.496 ug/l

RT: 9.234 min Scan# 1215

Delta R.T. 0.006 min

Lab File: VY017133.D

Acq: 30 Jan 2024 09:31

Instrument :

MSVOA_Y





ClientSampleId :

Tgt Ion: 63 Resp: 71195

Ion Ratio Lower Upper

63 100

65 30.4 24.6 36.8



#46

Dibromomethane

Concen: 48.099 ug/l

RT: 9.325 min Scan# 1231

Delta R.T. 0.012 min

Lab File: VY017133.D

Acq: 30 Jan 2024 09:31




Tgt Ion: 93 Resp: 35316


Ion Ratio Lower Upper

93 100

95 82.3 66.6 99.8

174 90.9 73.8 110.6









#47
Bromodichloromethane
Concen: 50.727 ug/l
RT: 9.514 min Scan# 1262
Delta R.T. 0.006 min
Lab File: VY017133.D
Acq: 30 Jan 2024 09:31




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion: 83 Resp: 94755
Ion Ratio Lower Upper
83 100
85 63.5 53.2 79.8
127 8.3 6.4 9.6



#48
Methyl methacrylate
Concen: 46.619 ug/l
RT: 9.313 min Scan# 1229
Delta R.T. 0.012 min
Lab File: VY017133.D
Acq: 30 Jan 2024 09:31

Tgt Ion: 41 Resp: 40300
Ion Ratio Lower Upper
41 100
69 91.4 72.1 108.1
39 54.6 42.8 64.2









#49
1,4-Dioxane
Concen: 1009.236 ug/l
RT: 9.319 min Scan# 11
Delta R.T. 0.006 min
Lab File: VY017133.D
Acq: 30 Jan 2024 09:31




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion: 88 Resp: 7288
Ion Ratio Lower Upper
88 100
43 33.7 28.2 42.4
58 69.4 52.2 78.2



#50
Toluene-d8
Concen: 56.921 ug/l
RT: 10.197 min Scan# 1374
Delta R.T. 0.006 min
Lab File: VY017133.D
Acq: 30 Jan 2024 09:31

Tgt Ion: 98 Resp: 253442
Ion Ratio Lower Upper
98 100
100 65.7 52.3 78.5









#51
4-Methyl-2-Pentanone
Concen: 235.609 ug/l
RT: 10.087 min Scan# 11
Delta R.T. 0.006 min
Lab File: VY017133.D
Acq: 30 Jan 2024 09:31




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion: 43 Resp: 195133
Ion Ratio Lower Upper
43 100
58 39.7 32.2 48.2



#52
Toluene
Concen: 51.417 ug/l
RT: 10.264 min Scan# 1385
Delta R.T. 0.012 min
Lab File: VY017133.D
Acq: 30 Jan 2024 09:31

Tgt Ion: 92 Resp: 175298
Ion Ratio Lower Upper
92 100
91 172.3 137.6 206.4









#53
t-1,3-Dichloropropene
Concen: 49.771 ug/l
RT: 10.484 min Scan# 1420
Delta R.T. 0.013 min
Lab File: VY017133.D
Acq: 30 Jan 2024 09:31




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion: 75 Resp: 84929
Ion Ratio Lower Upper
75 100
77 31.7 24.6 36.8



#54
cis-1,3-Dichloropropene
Concen: 50.470 ug/l
RT: 9.947 min Scan# 1333
Delta R.T. 0.006 min
Lab File: VY017133.D
Acq: 30 Jan 2024 09:31


Tgt Ion: 75 Resp: 105923
Ion Ratio Lower Upper
75 100
77 31.5 24.6 36.8
39 43.4 33.6 50.4








#55
1,1,2-Trichloroethane
Concen: 48.487 ug/l
RT: 10.660 min Scan# 1450
Delta R.T. 0.006 min
Lab File: VY017133.D
Acq: 30 Jan 2024 09:31

Instrument :
MSVOA_Y
ClientSampleId :



Tgt Ion: 97 Resp: 49757




Ion	Ratio	Lower	Upper
97	100		
83	87.2	69.4	104.2
85	55.5	43.9	65.9
99	61.5	48.6	73.0




#56
Ethyl methacrylate
Concen: 47.619 ug/l
RT: 10.526 min Scan# 1428
Delta R.T. 0.006 min
Lab File: VY017133.D
Acq: 30 Jan 2024 09:31

Tgt Ion: 69 Resp: 42733

Ion	Ratio	Lower	Upper
69	100		
41	65.0	49.2	73.8
39	32.6	26.8	40.2





#57

1,3-Dichloropropane

Concen: 49.542 ug/l

RT: 10.807 min Scan# 1473

Delta R.T. 0.007 min

Lab File: VY017133.D

Acq: 30 Jan 2024 09:31

Instrument :

MSVOA_Y





ClientSampleId :

Tgt Ion: 76 Resp: 85954

Ion Ratio Lower Upper

76 100

78 32.1 26.1 39.1



#58

2-Chloroethyl Vinyl ether

Concen: 295.642 ug/l

RT: 9.801 min Scan# 1309

Delta R.T. 0.013 min

Lab File: VY017133.D




Acq: 30 Jan 2024 09:31


Tgt Ion: 63 Resp: 184161

Ion Ratio Lower Upper

63 100

106 27.8 22.4 33.6









#59
2-Hexanone
Concen: 216.429 ug/l
RT: 10.849 min Scan# 1480
Delta R.T. 0.006 min
Lab File: VY017133.D
Acq: 30 Jan 2024 09:31




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion: 43 Resp: 136750
Ion Ratio Lower Upper
43 100
58 56.8 28.1 84.3



#60
Dibromochloromethane
Concen: 48.985 ug/l
RT: 11.002 min Scan# 1506
Delta R.T. 0.006 min
Lab File: VY017133.D
Acq: 30 Jan 2024 09:31

Tgt Ion:129 Resp: 58401
Ion Ratio Lower Upper
129 100
127 78.6 39.2 117.6









#61
1,2-Dibromoethane
Concen: 47.842 ug/l
RT: 11.105 min Scan# 11
Delta R.T. 0.006 min
Lab File: VY017133.D
Acq: 30 Jan 2024 09:31




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion:107 Resp: 44600
Ion Ratio Lower Upper
107 100
109 94.7 75.4 113.2



#62
4-Bromofluorobenzene
Concen: 55.750 ug/l
RT: 12.502 min Scan# 1752
Delta R.T. 0.013 min
Lab File: VY017133.D
Acq: 30 Jan 2024 09:31

Tgt Ion: 95 Resp: 76468
Ion Ratio Lower Upper
95 100
174 81.7 0.0 160.0
176 78.3 0.0 151.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.508 min Scan# 11
Delta R.T. 0.006 min
Lab File: VY017133.D
Acq: 30 Jan 2024 09:31




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion:117 Resp: 166967
Ion Ratio Lower Upper
117 100
82 53.2 43.8 65.8
119 31.0 26.5 39.7



#64
Tetrachloroethene
Concen: 48.929 ug/l
RT: 10.740 min Scan# 1463
Delta R.T. 0.013 min
Lab File: VY017133.D
Acq: 30 Jan 2024 09:31

Tgt Ion:164 Resp: 56944
Ion Ratio Lower Upper
164 100
166 134.5 98.8 148.2
129 93.8 78.9 118.3
131 91.0 72.5 108.7









#65
Chlorobenzene
Concen: 50.096 ug/l
RT: 11.532 min Scan# 11
Delta R.T. 0.006 min
Lab File: VY017133.D
Acq: 30 Jan 2024 09:31




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion:112 Resp: 177253
Ion Ratio Lower Upper
112 100
114 31.8 26.1 39.1



#66
1,1,1,2-Tetrachloroethane
Concen: 51.432 ug/l
RT: 11.605 min Scan# 1605
Delta R.T. 0.006 min
Lab File: VY017133.D
Acq: 30 Jan 2024 09:31


Tgt Ion:131 Resp: 66471
Ion Ratio Lower Upper
131 100
133 94.9 47.0 141.0
119 67.6 33.3 99.9








#67
Ethyl Benzene
Concen: 51.966 ug/l
RT: 11.611 min Scan# 1606
Delta R.T. 0.012 min
Lab File: VY017133.D
Acq: 30 Jan 2024 09:31

Instrument :
MSVOA_Y
ClientSampleId :







Tgt Ion: 91 Resp: 334257
Ion Ratio Lower Upper
91 100
106 29.8 25.0 37.4



#68
m/p-Xylenes
Concen: 105.388 ug/l
RT: 11.721 min Scan# 1624
Delta R.T. 0.012 min
Lab File: VY017133.D
Acq: 30 Jan 2024 09:31

Tgt Ion:106 Resp: 254690
Ion Ratio Lower Upper
106 100
91 198.0 160.7 241.1









#69
o-Xylene
Concen: 51.827 ug/l
RT: 12.044 min Scan# 1680
Delta R.T. 0.006 min
Lab File: VY017133.D
Acq: 30 Jan 2024 09:31




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion:106 Resp: 113772
Ion Ratio Lower Upper
106 100
91 213.8 106.5 319.5



#70
Styrene
Concen: 53.439 ug/l
RT: 12.063 min Scan# 1680
Delta R.T. 0.013 min
Lab File: VY017133.D
Acq: 30 Jan 2024 09:31

Tgt Ion:104 Resp: 170306
Ion Ratio Lower Upper
104 100
78 48.2 39.2 58.8
103 55.7 43.8 65.6









#71
Bromoform
Concen: 49.502 ug/l
RT: 12.227 min Scan# 1706
Delta R.T. 0.012 min
Lab File: VY017133.D
Acq: 30 Jan 2024 09:31




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion:173 Resp: 32080
Ion Ratio Lower Upper
173 100
175 48.3 23.5 70.6
254 0.0 0.0 0.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.440 min Scan# 1906
Delta R.T. 0.006 min
Lab File: VY017133.D
Acq: 30 Jan 2024 09:31

Tgt Ion:152 Resp: 72856
Ion Ratio Lower Upper
152 100
115 58.3 30.0 90.0
150 175.3 0.0 344.6









#73
Isopropylbenzene
Concen: 53.167 ug/l
RT: 12.349 min Scan# 1727
Delta R.T. 0.012 min
Lab File: VY017133.D
Acq: 30 Jan 2024 09:31




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion:105 Resp: 318437
Ion Ratio Lower Upper
105 100
120 26.1 13.0 39.0



#74
N-ethyl acetate
Concen: 46.944 ug/l
RT: 12.160 min Scan# 1696
Delta R.T. 0.006 min
Lab File: VY017133.D
Acq: 30 Jan 2024 09:31

Tgt Ion: 43 Resp: 56673
Ion Ratio Lower Upper
43 100
70 46.1 36.6 55.0
55 27.6 22.2 33.4
61 24.7 20.2 30.4









#75
1,1,2,2-Tetrachloroethane
Concen: 47.631 ug/l
RT: 12.599 min Scan# 1767
Delta R.T. 0.006 min
Lab File: VY017133.D
Acq: 30 Jan 2024 09:31




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion: 83 Resp: 54399
Ion Ratio Lower Upper
83 100
131 9.1 4.9 14.6
85 64.8 32.6 98.0



#76
1,2,3-Trichloropropane
Concen: 75.081 ug/l
RT: 12.648 min Scan# 1776
Delta R.T. 0.006 min
Lab File: VY017133.D
Acq: 30 Jan 2024 09:31

Tgt Ion: 75 Resp: 66396
Ion Ratio Lower Upper
75 100
77 0.0 0.0 0.0









#77
Bromobenzene
Concen: 51.303 ug/l
RT: 12.623 min Scan# 1772
Delta R.T. 0.006 min
Lab File: VY017133.D
Acq: 30 Jan 2024 09:31




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion:156 Resp: 66487
Ion Ratio Lower Upper
156 100
77 188.6 97.3 291.8
158 96.5 48.4 145.0



#78
n-propylbenzene
Concen: 53.324 ug/l
RT: 12.691 min Scan# 1783
Delta R.T. 0.013 min
Lab File: VY017133.D
Acq: 30 Jan 2024 09:31

Tgt Ion: 91 Resp: 384134
Ion Ratio Lower Upper
91 100
120 22.0 11.0 33.0





#79

2-Chlorotoluene

Concen: 52.452 ug/l

RT: 12.776 min Scan# 1

Delta R.T. 0.013 min

Lab File: VY017133.D

Acq: 30 Jan 2024 09:31

Instrument :

MSVOA_Y





ClientSampleId :

Tgt Ion: 91 Resp: 207350

Ion Ratio Lower Upper

91 100

126 35.1 17.0 51.0



#80

1,3,5-Trimethylbenzene

Concen: 52.953 ug/l

RT: 12.831 min Scan# 1806

Delta R.T. 0.013 min

Lab File: VY017133.D




Acq: 30 Jan 2024 09:31


Tgt Ion:105 Resp: 251005

Ion Ratio Lower Upper

105 100

120 48.8 24.3 72.8









#81
trans-1,4-Dichloro-2-butene
Concen: 47.034 ug/l
RT: 12.398 min Scan# 11
Delta R.T. 0.013 min
Lab File: VY017133.D
Acq: 30 Jan 2024 09:31




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion: 75 Resp: 16066
Ion Ratio Lower Upper
75 100
53 90.4 74.0 111.0
89 44.8 36.5 54.7



#82
4-Chlorotoluene
Concen: 52.069 ug/l
RT: 12.873 min Scan# 1813
Delta R.T. 0.012 min
Lab File: VY017133.D
Acq: 30 Jan 2024 09:31

Tgt Ion: 91 Resp: 213330
Ion Ratio Lower Upper
91 100
126 34.0 17.1 51.2









#83
tert-Butylbenzene
Concen: 53.752 ug/l
RT: 13.093 min Scan# 1849
Delta R.T. 0.006 min
Lab File: VY017133.D
Acq: 30 Jan 2024 09:31




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion:119 Resp: 220389
Ion Ratio Lower Upper
119 100
91 63.2 31.5 94.5
134 26.9 13.1 39.3



#84
1,2,4-Trimethylbenzene
Concen: 52.520 ug/l
RT: 13.136 min Scan# 1856
Delta R.T. 0.007 min
Lab File: VY017133.D
Acq: 30 Jan 2024 09:31

Tgt Ion:105 Resp: 247437
Ion Ratio Lower Upper
105 100
120 45.3 22.1 66.3









#85
sec-Butylbenzene
Concen: 53.559 ug/l
RT: 13.270 min Scan# 1878
Delta R.T. 0.007 min
Lab File: VY017133.D
Acq: 30 Jan 2024 09:31




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion:105 Resp: 325159
Ion Ratio Lower Upper
105 100
134 19.7 9.7 28.9



#86
p-Isopropyltoluene
Concen: 53.666 ug/l
RT: 13.386 min Scan# 1897
Delta R.T. 0.007 min
Lab File: VY017133.D
Acq: 30 Jan 2024 09:31

Tgt Ion:119 Resp: 271394
Ion Ratio Lower Upper
119 100
134 26.4 13.1 39.1
91 23.1 11.7 35.0





#87

1,3-Dichlorobenzene

Concen: 51.307 ug/l

RT: 13.386 min Scan# 1910

Delta R.T. 0.013 min

Lab File: VY017133.D

Acq: 30 Jan 2024 09:31

Instrument :

MSVOA_Y

ClientSampleId :





Tgt Ion:146 Resp: 131279

Ion Ratio Lower Upper

146 100

111 40.0 20.2 60.5

148 63.4 31.9 95.9



#88

1,4-Dichlorobenzene

Concen: 49.583 ug/l

RT: 13.465 min Scan# 1910

Delta R.T. 0.013 min

Lab File: VY017133.D

Acq: 30 Jan 2024 09:31




Tgt Ion:146 Resp: 124824


Ion Ratio Lower Upper

146 100

111 39.0 19.3 57.9

148 64.3 31.6 94.7









#89
n-Butylbenzene
Concen: 53.454 ug/l
RT: 13.715 min Scan# 1951
Delta R.T. 0.013 min
Lab File: VY017133.D
Acq: 30 Jan 2024 09:31




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion: 91 Resp: 260743
Ion Ratio Lower Upper
91 100
92 52.1 26.4 79.2
134 24.9 12.3 36.8



#90
Hexachloroethane
Concen: 51.918 ug/l
RT: 13.977 min Scan# 1994
Delta R.T. 0.006 min
Lab File: VY017133.D
Acq: 30 Jan 2024 09:31

Tgt Ion: 117 Resp: 54724
Ion Ratio Lower Upper
117 100
201 66.9 33.4 100.1





#91

1,2-Dichlorobenzene

Concen: 50.330 ug/l

RT: 13.757 min Scan# 1957

Delta R.T. 0.006 min

Lab File: VY017133.D

Acq: 30 Jan 2024 09:31

Instrument :

MSVOA_Y

ClientSampleId :





Tgt Ion:146 Resp: 109975

Ion Ratio Lower Upper

146 100

111 41.1 20.5 61.6

148 64.7 31.4 94.3



#92

1,2-Dibromo-3-Chloropropane

Concen: 43.883 ug/l

RT: 14.373 min Scan# 2059

Delta R.T. 0.006 min

Lab File: VY017133.D

Acq: 30 Jan 2024 09:31




Tgt Ion: 75 Resp: 6401


Ion Ratio Lower Upper

75 100

155 85.0 42.1 126.3


157 113.6 55.8 167.4








#93
1,2,4-Trichlorobenzene
Concen: 48.852 ug/l
RT: 15.025 min Scan# 2166
Delta R.T. 0.006 min
Lab File: VY017133.D
Acq: 30 Jan 2024 09:31


Instrument :
MSVOA_Y
ClientSampleId :





Tgt Ion:180 Resp: 55782
Ion Ratio Lower Upper
180 100
182 94.2 47.9 143.7
145 30.0 15.9 47.7




#94
Hexachlorobutadiene
Concen: 48.906 ug/l
RT: 15.129 min Scan# 2183
Delta R.T. 0.006 min
Lab File: VY017133.D
Acq: 30 Jan 2024 09:31



Tgt Ion:225 Resp: 33609
Ion Ratio Lower Upper
225 100
223 63.6 31.4 94.3
227 65.4 32.3 96.8









#95
Naphthalene
Concen: 47.014 ug/l
RT: 15.257 min Scan# 21
Delta R.T. 0.012 min
Lab File: VY017133.D
Acq: 30 Jan 2024 09:31

Instrument :
MSVOA_Y
ClientSampleId :

Tgt Ion:128	Resp:	94815
Ion Ratio	Lower	Upper
128	100	
127	12.8	10.2 15.4
129	10.9	8.6 12.8



#96
1,2,3-Trichlorobenzene
Concen: 47.405 ug/l
RT: 15.446 min Scan# 2235
Delta R.T. 0.012 min
Lab File: VY017133.D
Acq: 30 Jan 2024 09:31

Tgt Ion:180	Resp:	45034
Ion Ratio	Lower	Upper
180	100	
182	95.4	46.3 138.8
145	32.6	16.1 48.2

