

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY013024\
 Data File : VY017137.D
 Acq On : 30 Jan 2024 12:41
 Operator : SY/MD
 Sample : P1276-03
 Misc : 2.56g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP-37

Quant Time: Jan 31 00:27:24 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y012424S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 25 04:44:10 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.813	168	91248	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	8.710	114	191794	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.508	117	174469	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.446	152	53105	50.000	ug/l	0.01
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.161	65	40828	45.924	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	91.840%
35) Dibromofluoromethane	7.740	113	46061	39.791	ug/l	0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	79.580%
50) Toluene-d8	10.197	98	169779	38.832	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	77.660%
62) 4-Bromofluorobenzene	12.502	95	51984	38.596	ug/l	0.01
Spiked Amount	50.000	Range	30 - 143	Recovery	=	77.200%
Target Compounds						
73) Isopropylbenzene	12.355	105	8940	2.048	ug/l	Qvalue # 70

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY013024\
Data File : VY017137.D
Acq On : 30 Jan 2024 12:41
Operator : SY/MD
Sample : P1276-03
Misc : 2.56g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP-37

Quant Time: Jan 31 00:27:24 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y012424S.M
Quant Title : SW846 8260
QLast Update : Thu Jan 25 04:44:10 2024
Response via : Initial Calibration

