

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY013024\
 Data File : VY017143.D
 Acq On : 30 Jan 2024 15:02
 Operator : SY/MD
 Sample : P1285-01
 Misc : 5.05g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SECOND-AVE-TP-1

Quant Time: Jan 31 00:28:19 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y012424S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 25 04:44:10 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.807	168	15984	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.710	114	28146	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.502	117	25173	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.440	152	7953	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.161	65	7172	46.053	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	92.100%
35) Dibromofluoromethane	7.740	113	7620	44.857	ug/l	0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	89.720%
50) Toluene-d8	10.197	98	24129	37.607	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	75.220%
62) 4-Bromofluorobenzene	12.496	95	6615	33.468	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	66.940%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY013024\
Data File : VY017143.D
Acq On : 30 Jan 2024 15:02
Operator : SY/MD
Sample : P1285-01
Misc : 5.05g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SECOND-AVE-TP-1

Quant Time: Jan 31 00:28:19 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y012424S.M
Quant Title : SW846 8260
QLast Update : Thu Jan 25 04:44:10 2024
Response via : Initial Calibration

