

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY013024\
 Data File : VY017147.D
 Acq On : 30 Jan 2024 16:36
 Operator : SY/MD
 Sample : P1285-01RE
 Misc : 5.08g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SECOND-AVE-TP-1

Quant Time: Jan 31 00:29:04 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y012424S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 25 04:44:10 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	78157	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.704	114	166747	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.502	117	170150	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	69281	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.155	65	38376	50.396	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	100.800%
35) Dibromofluoromethane	7.728	113	41247	40.985	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	81.980%
50) Toluene-d8	10.191	98	151243	39.789	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	79.580%
62) 4-Bromofluorobenzene	12.489	95	54299	46.371	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	92.740%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY013024\
Data File : VY017147.D
Acq On : 30 Jan 2024 16:36
Operator : SY/MD
Sample : P1285-01RE
Misc : 5.08g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SECOND-AVE-TP-1

Quant Time: Jan 31 00:29:04 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y012424S.M
Quant Title : SW846 8260
QLast Update : Thu Jan 25 04:44:10 2024
Response via : Initial Calibration

