



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY013024\
Data File : VY017132.D
Acq On : 30 Jan 2024 08:57
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y012424S.M
Title : SW846 8260
Last Update : Thu Jan 25 04:44:10 2024



AutoFind: Scans 1751, 1752, 1753; Background Corrected with Scan 1743

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	19.6	10986	PASS
75	95	30	60	52.5	29395	PASS
95	95	100	100	100.0	55997	PASS
96	95	5	9	6.7	3742	PASS
173	174	0.00	2	1.4	615	PASS
174	95	50	100	81.1	45405	PASS
175	174	5	9	7.3	3297	PASS
176	174	95	101	97.5	44264	PASS
177	176	5	9	5.8	2584	PASS