

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY013025\
 Data File : VY020986.D
 Acq On : 30 Jan 2025 15:26
 Operator : SY/MD
 Sample : Q1206-05
 Misc : 5.10g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 JPP-16.3-012725

Quant Time: Jan 31 04:41:35 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y013025S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jan 31 02:19:37 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	334296	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	513886	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	376362	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.353	152	132171	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.067	65	120643	42.387	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	84.780%
35) Dibromofluoromethane	7.640	113	130327	42.185	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	84.380%
50) Toluene-d8	10.109	98	458890	38.057	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	76.120%
62) 4-Bromofluorobenzene	12.414	95	122132	30.185	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	60.360%

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY013025\
Data File : VY020986.D
Acq On : 30 Jan 2025 15:26
Operator : SY/MD
Sample : Q1206-05
Misc : 5.10g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
JPP-16.3-012725

Quant Time: Jan 31 04:41:35 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y013025S.M
Quant Title : SW846 8260
QLast Update : Fri Jan 31 02:19:37 2025
Response via : Initial Calibration

