

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY013025\
 Data File : VY020987.D
 Acq On : 30 Jan 2025 15:49
 Operator : SY/MD
 Sample : Q1207-01
 Misc : 4.99g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 JPP-2.1-012725

Quant Time: Jan 31 04:41:54 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y013025S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jan 31 02:19:37 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	318303	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	481883	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	373394	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.353	152	146856	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.067	65	113973	42.055	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	84.120%
35) Dibromofluoromethane	7.640	113	125069	43.172	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	86.340%
50) Toluene-d8	10.109	98	455466	40.281	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	80.560%
62) 4-Bromofluorobenzene	12.408	95	130832	34.483	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	68.960%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY013025\
Data File : VY020987.D
Acq On : 30 Jan 2025 15:49
Operator : SY/MD
Sample : Q1207-01
Misc : 4.99g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
JPP-2.1-012725

Quant Time: Jan 31 04:41:54 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y013025S.M
Quant Title : SW846 8260
QLast Update : Fri Jan 31 02:19:37 2025
Response via : Initial Calibration

