

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY013025\
 Data File : VY020993.D
 Acq On : 30 Jan 2025 18:05
 Operator : SY/MD
 Sample : Q1216-09
 Misc : 5.04g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 JPP-21.2-012825

Quant Time: Jan 31 04:43:46 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y013025S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jan 31 02:19:37 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	8991	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	13054	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	9470	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.347	152	1334	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	4665	60.940	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	121.880%
35) Dibromofluoromethane	7.634	113	4232	53.926	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	107.860%
50) Toluene-d8	10.109	98	12549	40.969	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	81.940%
62) 4-Bromofluorobenzene	12.414	95	2257	21.959	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	43.920%
Target Compounds						
5) Bromomethane	2.611	94	284	8.371	ug/l	Qvalue # 71

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY013025\
Data File : VY020993.D
Acq On : 30 Jan 2025 18:05
Operator : SY/MD
Sample : Q1216-09
Misc : 5.04g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
JPP-21.2-012825

Quant Time: Jan 31 04:43:46 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y013025S.M
Quant Title : SW846 8260
QLast Update : Fri Jan 31 02:19:37 2025
Response via : Initial Calibration

