

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY013123\
 Data File : VY012419.D
 Acq On : 31 Jan 2023 14:48
 Operator : KP/MD
 Sample : 01352-01
 Misc : 5.63g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP-2-WC

Quant Time: Feb 01 01:24:37 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011623S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 04:31:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	156122	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	261281	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	259762	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	124208	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	96438	76.616	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	153.240%
35) Dibromofluoromethane	7.716	113	78946	57.063	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	114.120%
50) Toluene-d8	10.179	98	313503	66.862	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	133.720%
62) 4-Bromofluorobenzene	12.483	95	105290	53.987	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	107.980%
Target Compounds						Qvalue
16) Acetone	3.948	43	14843	34.143	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY013123\
Data File : VY012419.D
Acq On : 31 Jan 2023 14:48
Operator : KP/MD
Sample : 01352-01
Misc : 5.63g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP-2-WC

Quant Time: Feb 01 01:24:37 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011623S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 17 04:31:47 2023
Response via : Initial Calibration

