

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY013123\
 Data File : VY012420.D
 Acq On : 31 Jan 2023 15:11
 Operator : KP/MD
 Sample : 01355-02RE
 Misc : 5.10g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VOCRE

Quant Time: Feb 01 01:24:49 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011623S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 04:31:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	20581	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	34260	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	33889	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	14208	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	13728	83.073	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 166.140%#	
35) Dibromofluoromethane	7.716	113	10990	60.582	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 121.160%	
50) Toluene-d8	10.179	98	40865	66.440	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 132.880%	
62) 4-Bromofluorobenzene	12.483	95	12340	48.254	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 96.500%	
Target Compounds						
16) Acetone	3.942	43	757	13.209	ug/l	Qvalue 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY013123\
Data File : VY012420.D
Acq On : 31 Jan 2023 15:11
Operator : KP/MD
Sample : 01355-02RE
Misc : 5.10g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VOCRE

Quant Time: Feb 01 01:24:49 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011623S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 17 04:31:47 2023
Response via : Initial Calibration

