

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY013123\
 Data File : VY012421.D
 Acq On : 31 Jan 2023 15:35
 Operator : KP/MD
 Sample : 01358-04RE
 Misc : 7.57g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 DR-2-012723RE

Quant Time: Feb 01 01:25:00 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011623S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 04:31:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	155094	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	258645	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	254188	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	117163	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	94649	75.642	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	151.280%
35) Dibromofluoromethane	7.716	113	76898	56.149	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	112.300%
50) Toluene-d8	10.179	98	312292	67.313	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	134.620%
62) 4-Bromofluorobenzene	12.483	95	101463	52.555	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	105.100%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY013123\
Data File : VY012421.D
Acq On : 31 Jan 2023 15:35
Operator : KP/MD
Sample : 01358-04RE
Misc : 7.57g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
DR-2-012723RE

Quant Time: Feb 01 01:25:00 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011623S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 17 04:31:47 2023
Response via : Initial Calibration

