

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY013123\
 Data File : VY012426.D
 Acq On : 31 Jan 2023 17:32
 Operator : KP/MD
 Sample : 01371-03
 Misc : 7.32g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 3-RT-15-S

Quant Time: Feb 01 01:26:04 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011623S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 04:31:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	217066	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	367795	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	356978	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	164513	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	128789	73.422	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	146.840%
35) Dibromofluoromethane	7.716	113	112304	57.666	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	115.340%
50) Toluene-d8	10.179	98	437062	66.174	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	132.340%
62) 4-Bromofluorobenzene	12.483	95	149243	54.362	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	108.720%
Target Compounds						
20) Methylene Chloride	4.716	84	37435	8.143	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY013123\
Data File : VY012426.D
Acq On : 31 Jan 2023 17:32
Operator : KP/MD
Sample : 01371-03
Misc : 7.32g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
3-RT-15-S

Quant Time: Feb 01 01:26:04 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011623S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 17 04:31:47 2023
Response via : Initial Calibration

