

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY013123\
 Data File : VY012428.D
 Acq On : 31 Jan 2023 18:19
 Operator : KP/MD
 Sample : 01366-01
 Misc : 4.99g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 NB-07-013023

Quant Time: Feb 01 01:26:27 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011623S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 04:31:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	211994	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	354856	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	346252	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	164165	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	114633	66.537	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	133.080%
35) Dibromofluoromethane	7.716	113	104912	55.835	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	111.680%
50) Toluene-d8	10.179	98	422185	66.258	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	132.520%
62) 4-Bromofluorobenzene	12.483	95	143229	54.074	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	108.140%
Target Compounds						
84) 1,2,4-Trimethylbenzene	13.123	105	12739	1.255	ug/l	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY013123\
Data File : VY012428.D
Acq On : 31 Jan 2023 18:19
Operator : KP/MD
Sample : 01366-01
Misc : 4.99g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
NB-07-013023

Quant Time: Feb 01 01:26:27 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011623S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 17 04:31:47 2023
Response via : Initial Calibration

