

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020123\
 Data File : VY012434.D
 Acq On : 01 Feb 2023 11:56
 Operator : KP/MD
 Sample : VY0201SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0201SBL01

Quant Time: Feb 01 23:02:26 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011623S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 04:31:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	168562	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	260736	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	234510	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	114018	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	71256	51.085	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	102.160%
35) Dibromofluoromethane	7.722	113	65195	47.222	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	94.440%
50) Toluene-d8	10.179	98	218555	45.280	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	90.560%
62) 4-Bromofluorobenzene	12.483	95	81721	41.990	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	83.980%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020123\
Data File : VY012434.D
Acq On : 01 Feb 2023 11:56
Operator : KP/MD
Sample : VY0201SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0201SBL01

Quant Time: Feb 01 23:02:26 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011623S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 17 04:31:47 2023
Response via : Initial Calibration

