

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020123\
 Data File : VY012445.D
 Acq On : 01 Feb 2023 16:26
 Operator : KP/MD
 Sample : 01386-13
 Misc : 6.45g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 JPP-15-013123

Quant Time: Feb 01 23:06:24 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011623S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 04:31:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	173002	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	291641	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	282836	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	129667	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	106235	76.139	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	152.280%
35) Dibromofluoromethane	7.716	113	83770	54.247	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	108.500%
50) Toluene-d8	10.179	98	348017	66.471	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	132.940%
62) 4-Bromofluorobenzene	12.483	95	116568	53.548	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	107.100%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020123\
Data File : VY012445.D
Acq On : 01 Feb 2023 16:26
Operator : KP/MD
Sample : 01386-13
Misc : 6.45g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
JPP-15-013123

Quant Time: Feb 01 23:06:24 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011623S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 17 04:31:47 2023
Response via : Initial Calibration

