

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020123\
 Data File : VY012449.D
 Acq On : 01 Feb 2023 18:00
 Operator : KP/MD
 Sample : 01401-01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 OK-1-212023

Quant Time: Feb 01 23:07:51 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011623S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 04:31:47 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	152022	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	262305	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	252212	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	112752	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	92953	75.796	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 151.600%	
35) Dibromofluoromethane	7.716	113	81245	58.496	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 117.000%	
50) Toluene-d8	10.179	98	306923	65.086	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 130.180%	
62) 4-Bromofluorobenzene	12.483	95	103525	52.875	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 105.740%	
Target Compounds						
						Qvalue
16) Acetone	3.948	43	6146	14.519	ug/l	90
52) Toluene	10.246	92	12440	2.742	ug/l	99
68) m/p-Xylenes	11.703	106	3969	1.088	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020123\
Data File : VY012449.D
Acq On : 01 Feb 2023 18:00
Operator : KP/MD
Sample : 01401-01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
OK-1-212023

Quant Time: Feb 01 23:07:51 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011623S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 17 04:31:47 2023
Response via : Initial Calibration

