



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020124\
Data File : VY017163.D
Acq On : 01 Feb 2024 09:15
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020124S.M
Title : SW846 8260
Last Update : Fri Feb 02 00:56:04 2024



AutoFind: Scans 1750, 1751, 1752; Background Corrected with Scan 1742

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	20.5	14436	PASS
75	95	30	60	53.0	37237	PASS
95	95	100	100	100.0	70280	PASS
96	95	5	9	6.7	4736	PASS
173	174	0.00	2	0.8	478	PASS
174	95	50	100	81.9	57563	PASS
175	174	5	9	7.1	4098	PASS
176	174	95	101	95.8	55120	PASS
177	176	5	9	6.4	3508	PASS