

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020223\
 Data File : VY012453.D
 Acq On : 02 Feb 2023 13:38
 Operator : KP/MD
 Sample : VY0202SBL07
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0202SBL07

Quant Time: Feb 03 03:11:12 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011623S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 04:31:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	170024	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	263914	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	235917	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	113276	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	69991	49.635	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	99.260%
35) Dibromofluoromethane	7.722	113	67881	48.576	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	97.160%
50) Toluene-d8	10.179	98	225835	46.324	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	92.640%
62) 4-Bromofluorobenzene	12.483	95	83284	42.277	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	84.560%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020223\
Data File : VY012453.D
Acq On : 02 Feb 2023 13:38
Operator : KP/MD
Sample : VY0202SBL07
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0202SBL07

Quant Time: Feb 03 03:11:12 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011623S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 17 04:31:47 2023
Response via : Initial Calibration

