

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020223\
 Data File : VY012451.D
 Acq On : 02 Feb 2023 11:58
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050

Manual Integrations
 APPROVED

Reviewed By :Krupa Patel 02/03/2023
 Supervised By :Mahesh Dadoda 02/03/2023

Quant Time: Feb 03 03:10:38 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011623S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 04:31:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	197045	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	302298	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	272875	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	134846	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	75875	46.153	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery =	92.300%		
35) Dibromofluoromethane	7.722	113	75470	47.149	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery =	94.300%		
50) Toluene-d8	10.185	98	255249	45.647	ug/l	0.00
Spiked Amount 50.000	Range 49 - 140		Recovery =	91.300%		
62) 4-Bromofluorobenzene	12.483	95	102352	45.360	ug/l	0.00
Spiked Amount 50.000	Range 25 - 144		Recovery =	90.720%		
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.906	85	69598	40.417	ug/l	96
3) Chloromethane	2.119	50	74742	47.018	ug/l	100
4) Vinyl Chloride	2.259	62	90720	48.893	ug/l	97
5) Bromomethane	2.656	94	67104	46.777	ug/l	97
6) Chloroethane	2.802	64	61046	50.447	ug/l	99
7) Trichlorofluoromethane	3.131	101	175939	48.584	ug/l	97
8) Diethyl Ether	3.533	74	56994	52.891	ug/l	87
9) 1,1,2-Trichlorotrifluo...	3.905	101	95990	54.463	ug/l	96
10) Methyl Iodide	4.094	142	130258	48.632	ug/l	99
11) Tert butyl alcohol	4.948	59	57862	301.046	ug/l	97
12) 1,1-Dichloroethene	3.881	96	93506	51.668	ug/l	97
13) Acrolein	3.728	56	19908	278.830	ug/l	99
14) Allyl chloride	4.484	41	128139	57.264	ug/l	93
15) Acrylonitrile	5.167	53	137483	292.093	ug/l	99
16) Acetone	3.948	43	134486	245.107	ug/l	100
17) Carbon Disulfide	4.204	76	252502	45.108	ug/l	99
18) Methyl Acetate	4.478	43	71063	57.880	ug/l	91
19) Methyl tert-butyl Ether	5.228	73	275870	54.684	ug/l	96
20) Methylene Chloride	4.722	84	113050	49.051	ug/l	91
21) trans-1,2-Dichloroethene	5.222	96	103309	51.156	ug/l	91
22) Diisopropyl ether	6.124	45	266521	59.743	ug/l #	96
23) Vinyl Acetate	6.063	43	813693	292.709	ug/l	95
24) 1,1-Dichloroethane	6.021	63	172588	55.776	ug/l	99
25) 2-Butanone	6.990	43	167795	277.010	ug/l	91
26) 2,2-Dichloropropane	6.984	77	173118	51.538	ug/l	99
27) cis-1,2-Dichloroethene	6.990	96	120013	54.101	ug/l	95
28) Bromochloromethane	7.338	49	33773	49.976	ug/l	83
29) Tetrahydrofuran	7.350	42	99522	299.111	ug/l	88
30) Chloroform	7.508	83	194318	53.199	ug/l	98
31) Cyclohexane	7.789	56	145813	50.754	ug/l	91
32) 1,1,1-Trichloroethane	7.703	97	185449	51.284	ug/l	98
36) 1,1-Dichloropropene	7.923	75	138463	50.899	ug/l	98
37) Ethyl Acetate	7.075	43	67065	55.923	ug/l #	94
38) Carbon Tetrachloride	7.905	117	172139	48.995	ug/l	98
39) Methylcyclohexane	9.185	83	173690	52.247	ug/l	97
40) Benzene	8.161	78	415565	52.927	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020223\
 Data File : VY012451.D
 Acq On : 02 Feb 2023 11:58
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050

Manual Integrations
 APPROVED

Reviewed By :Krupa Patel 02/03/2023
 Supervised By :Mahesh Dadoda 02/03/2023

Quant Time: Feb 03 03:10:38 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011623S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 04:31:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.325	41	38589m	54.743	ug/l	
42) 1,2-Dichloroethane	8.240	62	126269	50.515	ug/l	100
43) Isopropyl Acetate	8.270	43	128431	57.562	ug/l	95
44) Trichloroethene	8.941	130	122924	52.379	ug/l	97
45) 1,2-Dichloropropane	9.221	63	94759	57.838	ug/l	98
46) Dibromomethane	9.307	93	58050	52.927	ug/l	95
47) Bromodichloromethane	9.496	83	149227	54.050	ug/l	99
48) Methyl methacrylate	9.295	41	59006	57.397	ug/l	93
49) 1,4-Dioxane	9.301	88	16678	1115.943	ug/l #	93
51) 4-Methyl-2-Pentanone	10.069	43	345822	290.672	ug/l	95
52) Toluene	10.246	92	276786	52.945	ug/l	99
53) t-1,3-Dichloropropene	10.465	75	153900	53.291	ug/l	99
54) cis-1,3-Dichloropropene	9.929	75	169734	54.575	ug/l	99
55) 1,1,2-Trichloroethane	10.648	97	85600	54.932	ug/l	97
56) Ethyl methacrylate	10.514	69	116827	57.146	ug/l	94
57) 1,3-Dichloropropane	10.794	76	142201	55.876	ug/l	100
58) 2-Chloroethyl Vinyl ether	9.782	63	175675	259.699	ug/l	92
59) 2-Hexanone	10.831	43	248609	274.463	ug/l	94
60) Dibromochloromethane	10.989	129	111981	54.181	ug/l	99
61) 1,2-Dibromoethane	11.093	107	81463	54.176	ug/l	100
64) Tetrachloroethene	10.721	164	135826	54.550	ug/l	98
65) Chlorobenzene	11.520	112	298731	53.506	ug/l	99
66) 1,1,1,2-Tetrachloroethane	11.593	131	114836	54.492	ug/l	98
67) Ethyl Benzene	11.593	91	535197	53.637	ug/l	100
68) m/p-Xylenes	11.703	106	420660	106.625	ug/l	98
69) o-Xylene	12.032	106	202007	53.607	ug/l	98
70) Styrene	12.044	104	342341	54.373	ug/l	99
71) Bromoform	12.209	173	72173	53.604	ug/l #	98
73) Isopropylbenzene	12.331	105	542567	55.799	ug/l	99
74) N-amyl acetate	12.148	43	112506	59.024	ug/l	95
75) 1,1,2,2-Tetrachloroethane	12.587	83	90133	58.850	ug/l	100
76) 1,2,3-Trichloropropane	12.635	75	80060m	64.567	ug/l	
77) Bromobenzene	12.611	156	128546	54.939	ug/l	93
78) n-propylbenzene	12.672	91	634373	55.752	ug/l	99
79) 2-Chlorotoluene	12.757	91	354865	55.160	ug/l	98
80) 1,3,5-Trimethylbenzene	12.812	105	461193	54.560	ug/l	99
81) trans-1,4-Dichloro-2-b...	12.379	75	34259	56.928	ug/l	97
82) 4-Chlorotoluene	12.855	91	363543	53.797	ug/l	97
83) tert-Butylbenzene	13.080	119	408443	55.440	ug/l	99
84) 1,2,4-Trimethylbenzene	13.123	105	452658	54.301	ug/l	99
85) sec-Butylbenzene	13.257	105	584929	55.765	ug/l	98
86) p-Isopropyltoluene	13.373	119	497115	54.244	ug/l	99
87) 1,3-Dichlorobenzene	13.367	146	249292	53.570	ug/l	100
88) 1,4-Dichlorobenzene	13.446	146	247339	53.295	ug/l	99
89) n-Butylbenzene	13.696	91	441011	55.529	ug/l	99
90) Hexachloroethane	13.964	117	90891	54.938	ug/l	95
91) 1,2-Dichlorobenzene	13.745	146	224314	54.022	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	14.361	75	16867	53.289	ug/l	92
93) 1,2,4-Trichlorobenzene	15.007	180	136355	51.251	ug/l	100
94) Hexachlorobutadiene	15.111	225	81527	50.895	ug/l	99
95) Naphthalene	15.239	128	279968	53.313	ug/l	100
96) 1,2,3-Trichlorobenzene	15.428	180	118236	51.643	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020223\
Data File : VY012451.D
Acq On : 02 Feb 2023 11:58
Operator : KP/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050


Manual Integrations
APPROVED

Reviewed By :Krupa Patel 02/03/2023
Supervised By :Mahesh Dadoda 02/03/2023

Quant Time: Feb 03 03:10:38 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011623S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 17 04:31:47 2023
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed







#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.795 min Scan# 91
Delta R.T. 0.006 min
Lab File: VY012451.D
Acq: 02 Feb 2023 11:58

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion:168 Resp: 197045
Ion Ratio Lower Upper
168 100
99 55.9 44.0 66.0




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 02/03/2023
Supervised By :Mahesh Dadoda 02/03/2023



#2
Dichlorodifluoromethane
Concen: 40.417 ug/l
RT: 1.906 min Scan# 14
Delta R.T. -0.000 min
Lab File: VY012451.D
Acq: 02 Feb 2023 11:58


Tgt Ion: 85 Resp: 69598
Ion Ratio Lower Upper
85 100
87 31.6 16.8 50.4





#3
Chloromethane
Concen: 47.018 ug/l
RT: 2.119 min Scan# 49
Delta R.T. 0.006 min
Lab File: VY012451.D
Acq: 02 Feb 2023 11:58




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 50 Resp: 7474
Ion Ratio Lower Upper
50 100
52 33.6 26.8 40.2




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 02/03/2023
Supervised By :Mahesh Dadoda 02/03/2023



#4
Vinyl Chloride
Concen: 48.893 ug/l
RT: 2.259 min Scan# 72
Delta R.T. 0.006 min
Lab File: VY012451.D
Acq: 02 Feb 2023 11:58


Tgt Ion: 62 Resp: 90720
Ion Ratio Lower Upper
62 100
64 32.9 25.0 37.6





#5
Bromomethane
Concen: 46.777 ug/l
RT: 2.656 min Scan# 11
Delta R.T. 0.006 min
Lab File: VY012451.D
Acq: 02 Feb 2023 11:58




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion: 94 Resp: 67104
Ion Ratio Lower Upper
94 100
96 93.8 77.4 116.2

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 02/03/2023
Supervised By :Mahesh Dadoda 02/03/2023



#6
Chloroethane
Concen: 50.447 ug/l
RT: 2.802 min Scan# 161
Delta R.T. 0.006 min
Lab File: VY012451.D
Acq: 02 Feb 2023 11:58



Tgt Ion: 64 Resp: 61046
Ion Ratio Lower Upper
64 100
66 31.7 25.0 37.4









#7
Trichlorofluoromethane
Concen: 48.584 ug/l
RT: 3.131 min Scan# 214
Delta R.T. 0.006 min
Lab File: VY012451.D
Acq: 02 Feb 2023 11:58

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion:101 Resp: 175939
Ion Ratio Lower Upper
101 100
103 66.0 51.0 76.6




Manual Integrations
APPROVED

Reviewed By :Krupa Patel 02/03/2023
Supervised By :Mahesh Dadoda 02/03/2023



#8
Diethyl Ether
Concen: 52.891 ug/l
RT: 3.533 min Scan# 281
Delta R.T. 0.005 min
Lab File: VY012451.D
Acq: 02 Feb 2023 11:58

Tgt Ion: 74 Resp: 56994
Ion Ratio Lower Upper
74 100
45 81.2 46.8 140.3









#9
1,1,2-Trichlorotrifluoroethane
Concen: 54.463 ug/l
RT: 3.905 min Scan# 341
Delta R.T. 0.005 min
Lab File: VY012451.D
Acq: 02 Feb 2023 11:58

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion:101 Resp: 95990
Ion Ratio Lower Upper
101 100
85 45.1 36.2 54.2
151 85.1 64.2 96.2




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 02/03/2023
Supervised By :Mahesh Dadoda 02/03/2023



#10
Methyl Iodide
Concen: 48.632 ug/l
RT: 4.094 min Scan# 373
Delta R.T. -0.001 min
Lab File: VY012451.D
Acq: 02 Feb 2023 11:58

Tgt Ion:142 Resp: 130258
Ion Ratio Lower Upper
142 100
127 47.3 37.0 55.4
141 14.5 11.4 17.0









#11
Tert butyl alcohol
Concen: 301.046 ug/l
RT: 4.948 min Scan# 514
Delta R.T. -0.000 min
Lab File: VY012451.D
Acq: 02 Feb 2023 11:58

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion: 59 Resp: 57862
Ion Ratio Lower Upper
59 100
57 8.3 7.4 11.2




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 02/03/2023
Supervised By :Mahesh Dadoda 02/03/2023



#12
1,1-Dichloroethene
Concen: 51.668 ug/l
RT: 3.881 min Scan# 338
Delta R.T. 0.006 min
Lab File: VY012451.D
Acq: 02 Feb 2023 11:58


Tgt Ion: 96 Resp: 93506
Ion Ratio Lower Upper
96 100
61 149.3 122.0 183.0
98 64.8 48.6 73.0





#13
Acrolein
Concen: 278.830 ug/l
RT: 3.728 min Scan# 313
Delta R.T. -0.001 min
Lab File: VY012451.D
Acq: 02 Feb 2023 11:58




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 56 Resp: 19908
Ion Ratio Lower Upper
56 100
55 69.8 56.6 85.0




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 02/03/2023
Supervised By :Mahesh Dadoda 02/03/2023



#14
Allyl chloride
Concen: 57.264 ug/l
RT: 4.484 min Scan# 437
Delta R.T. 0.005 min
Lab File: VY012451.D
Acq: 02 Feb 2023 11:58


Tgt Ion: 41 Resp: 128139
Ion Ratio Lower Upper
41 100
39 70.4 53.5 80.3
76 45.6 31.1 46.7





#15
Acrylonitrile
Concen: 292.093 ug/l
RT: 5.167 min Scan# 548
Delta R.T. 0.006 min
Lab File: VY012451.D
Acq: 02 Feb 2023 11:58




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion: 53 Resp: 137483
Ion Ratio Lower Upper
53 100
52 80.9 65.1 97.7
51 36.2 28.7 43.1

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 02/03/2023
Supervised By :Mahesh Dadoda 02/03/2023




#16
Acetone
Concen: 245.107 ug/l
RT: 3.948 min Scan# 349
Delta R.T. 0.006 min
Lab File: VY012451.D
Acq: 02 Feb 2023 11:58



Tgt Ion: 43 Resp: 134486
Ion Ratio Lower Upper
43 100
58 34.7 27.9 41.9









#17
Carbon Disulfide
Concen: 45.108 ug/l
RT: 4.204 min Scan# 391
Delta R.T. 0.006 min
Lab File: VY012451.D
Acq: 02 Feb 2023 11:58

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion: 76 Resp: 25250
Ion Ratio Lower Upper
76 100
78 9.2 7.2 10.8




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 02/03/2023
Supervised By :Mahesh Dadoda 02/03/2023



#18
Methyl Acetate
Concen: 57.880 ug/l
RT: 4.478 min Scan# 436
Delta R.T. 0.005 min
Lab File: VY012451.D
Acq: 02 Feb 2023 11:58

Tgt Ion: 43 Resp: 71063
Ion Ratio Lower Upper
43 100
74 29.6 20.1 30.1









#19
Methyl tert-butyl Ether
Concen: 54.684 ug/l
RT: 5.228 min Scan# 51
Delta R.T. 0.006 min
Lab File: VY012451.D
Acq: 02 Feb 2023 11:58

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion: 73 Resp: 275870
Ion Ratio Lower Upper
73 100
57 19.8 17.4 26.0




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 02/03/2023
Supervised By :Mahesh Dadoda 02/03/2023



#20
Methylene Chloride
Concen: 49.051 ug/l
RT: 4.722 min Scan# 476
Delta R.T. 0.006 min
Lab File: VY012451.D
Acq: 02 Feb 2023 11:58


Tgt Ion: 84 Resp: 113050
Ion Ratio Lower Upper
84 100
49 105.4 95.4 143.2
51 31.7 29.9 44.9
86 64.3 49.6 74.4





#21
trans-1,2-Dichloroethene
Concen: 51.156 ug/l
RT: 5.222 min Scan# 511
Delta R.T. 0.000 min
Lab File: VY012451.D
Acq: 02 Feb 2023 11:58




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion: 96 Resp: 103309
Ion Ratio Lower Upper
96 100
61 123.0 108.0 162.0
98 60.4 51.6 77.4

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 02/03/2023
Supervised By :Mahesh Dadoda 02/03/2023




#22
Diisopropyl ether
Concen: 59.743 ug/l
RT: 6.124 min Scan# 706
Delta R.T. 0.005 min
Lab File: VY012451.D
Acq: 02 Feb 2023 11:58



Tgt Ion: 45 Resp: 266521
Ion Ratio Lower Upper
45 100
43 56.4 45.0 67.4
87 34.1 22.6 33.8#
59 13.1 9.4 14.0









#23
 Vinyl Acetate
 Concen: 292.709 ug/l
 RT: 6.063 min Scan# 696
 Delta R.T. 0.005 min
 Lab File: VY012451.D
 Acq: 02 Feb 2023 11:58

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050

Tgt Ion: 43 Resp: 81369
 Ion Ratio Lower Upper
 43 100
 86 13.2 9.0 13.4




Manual Integrations
 APPROVED


Reviewed By :Krupa Patel 02/03/2023
 Supervised By :Mahesh Dadoda 02/03/2023



#24
 1,1-Dichloroethane
 Concen: 55.776 ug/l
 RT: 6.021 min Scan# 689
 Delta R.T. 0.006 min
 Lab File: VY012451.D
 Acq: 02 Feb 2023 11:58

Tgt Ion: 63 Resp: 172588
 Ion Ratio Lower Upper
 63 100
 98 7.5 3.5 10.5
 100 4.6 2.2 6.6





#25

2-Butanone

Concen: 277.010 ug/l

RT: 6.990 min Scan# 847

Delta R.T. 0.006 min

Lab File: VY012451.D

Acq: 02 Feb 2023 11:58

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion: 43 Resp: 16779

Ion Ratio Lower Upper

43 100





72 32.9 22.6 34.0

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 02/03/2023

Supervised By :Mahesh Dadoda 02/03/2023



#26

2,2-Dichloropropane

Concen: 51.538 ug/l

RT: 6.984 min Scan# 847

Delta R.T. 0.006 min

Lab File: VY012451.D




Acq: 02 Feb 2023 11:58


Tgt Ion: 77 Resp: 173118

Ion Ratio Lower Upper

77 100


97 23.1 11.4 34.1





#27
cis-1,2-Dichloroethene
Concen: 54.101 ug/l
RT: 6.990 min Scan# 848
Delta R.T. 0.006 min
Lab File: VY012451.D
Acq: 02 Feb 2023 11:58




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 96 Resp: 12001
Ion Ratio Lower Upper
96 100
61 133.3 0.0 283.4
98 64.8 0.0 127.6




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 02/03/2023
Supervised By :Mahesh Dadoda 02/03/2023



#28
Bromochloromethane
Concen: 49.976 ug/l
RT: 7.338 min Scan# 905
Delta R.T. 0.006 min
Lab File: VY012451.D
Acq: 02 Feb 2023 11:58


Tgt Ion: 49 Resp: 33773
Ion Ratio Lower Upper
49 100
129 2.4 0.0 4.6
130 98.9 66.7 100.1





#29
Tetrahydrofuran
Concen: 299.111 ug/l
RT: 7.350 min Scan# 906
Delta R.T. 0.006 min
Lab File: VY012451.D
Acq: 02 Feb 2023 11:58




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 42 Resp: 9952
Ion Ratio Lower Upper
42 100
72 54.5 37.0 55.6
71 51.1 34.9 52.3




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 02/03/2023
Supervised By :Mahesh Dadoda 02/03/2023



#30
Chloroform
Concen: 53.199 ug/l
RT: 7.508 min Scan# 933
Delta R.T. 0.005 min
Lab File: VY012451.D
Acq: 02 Feb 2023 11:58


Tgt Ion: 83 Resp: 194318
Ion Ratio Lower Upper
83 100
85 62.8 51.5 77.3





#31
Cyclohexane
Concen: 50.754 ug/l
RT: 7.789 min Scan# 91
Delta R.T. 0.006 min
Lab File: VY012451.D
Acq: 02 Feb 2023 11:58




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion: 56 Resp: 14581
Ion Ratio Lower Upper
56 100
69 37.2 29.2 43.8
84 102.4 72.6 108.8

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 02/03/2023
Supervised By :Mahesh Dadoda 02/03/2023




#32
1,1,1-Trichloroethane
Concen: 51.284 ug/l
RT: 7.703 min Scan# 965
Delta R.T. 0.005 min
Lab File: VY012451.D
Acq: 02 Feb 2023 11:58



Tgt Ion: 97 Resp: 185449
Ion Ratio Lower Upper
97 100
99 64.4 51.0 76.4
61 38.9 33.0 49.4





#33

1,2-Dichloroethane-d4

Concen: 46.153 ug/l

RT: 8.142 min Scan# 1037

Delta R.T. -0.001 min

Lab File: VY012451.D

Acq: 02 Feb 2023 11:58

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion: 65 Resp: 75875

Ion Ratio Lower Upper

65 100





67 51.7 0.0 103.2

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 02/03/2023

Supervised By :Mahesh Dadoda 02/03/2023



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.691 min Scan# 1127

Delta R.T. 0.000 min

Lab File: VY012451.D

Acq: 02 Feb 2023 11:58




Tgt Ion:114 Resp: 302298


Ion Ratio Lower Upper

114 100

63 17.6 0.0 39.6


88 15.0 0.0 32.2





#35
Dibromofluoromethane
Concen: 47.149 ug/l
RT: 7.722 min Scan# 90
Delta R.T. 0.006 min
Lab File: VY012451.D
Acq: 02 Feb 2023 11:58




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion:113 Resp: 75470
Ion Ratio Lower Upper
113 100
111 101.7 83.0 124.4
192 21.3 16.0 24.0

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 02/03/2023
Supervised By :Mahesh Dadoda 02/03/2023




#36
1,1-Dichloropropene
Concen: 50.899 ug/l
RT: 7.923 min Scan# 1001
Delta R.T. 0.006 min
Lab File: VY012451.D
Acq: 02 Feb 2023 11:58



Tgt Ion: 75 Resp: 138463
Ion Ratio Lower Upper
75 100
110 38.4 18.1 54.1
77 31.4 25.1 37.7





#37

Ethyl Acetate

Concen: 55.923 ug/l

RT: 7.075 min Scan# 80

Delta R.T. -0.001 min

Lab File: VY012451.D

Acq: 02 Feb 2023 11:58

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion: 43 Resp: 6706





Ion	Ratio	Lower	Upper
43	100		
61	16.4	11.3	16.9
70	12.9	8.2	12.4

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 02/03/2023

Supervised By :Mahesh Dadoda 02/03/2023



#38

Carbon Tetrachloride

Concen: 48.995 ug/l

RT: 7.905 min Scan# 998




Delta R.T. 0.006 min


Lab File: VY012451.D

Acq: 02 Feb 2023 11:58

Tgt Ion:117 Resp: 172139

Ion	Ratio	Lower	Upper
117	100		
119	93.6	73.8	110.8
121	30.4	25.2	37.8





#39

Methylcyclohexane

Concen: 52.247 ug/l

RT: 9.185 min Scan# 1

Delta R.T. -0.000 min

Lab File: VY012451.D

Acq: 02 Feb 2023 11:58

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion: 83 Resp: 173690

Ion Ratio Lower Upper

83 100

55 70.5 58.2 87.2





98 50.9 38.6 58.0

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 02/03/2023

Supervised By :Mahesh Dadoda 02/03/2023



#40

Benzene

Concen: 52.927 ug/l

RT: 8.161 min Scan# 1040

Delta R.T. -0.000 min

Lab File: VY012451.D




Acq: 02 Feb 2023 11:58


Tgt Ion: 78 Resp: 415565

Ion Ratio Lower Upper

78 100


77 23.5 18.6 27.8





#41
Methacrylonitrile
Concen: 54.743 ug/l m
RT: 7.325 min Scan# 903
Delta R.T. -0.001 min
Lab File: VY012451.D
Acq: 02 Feb 2023 11:58




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 41 Resp: 38589
Ion Ratio Lower Upper
41 100
39 58.3 0.0 0.0
67 0.0 0.0 0.0
52 0.0 0.0 0.0




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 02/03/2023
Supervised By :Mahesh Dadoda 02/03/2023



#42
1,2-Dichloroethane
Concen: 50.515 ug/l
RT: 8.240 min Scan# 1053
Delta R.T. 0.006 min
Lab File: VY012451.D
Acq: 02 Feb 2023 11:58


Tgt Ion: 62 Resp: 126269
Ion Ratio Lower Upper
62 100
98 10.0 0.0 19.6





#43
Isopropyl Acetate
Concen: 57.562 ug/l
RT: 8.270 min Scan# 1058
Delta R.T. -0.001 min
Lab File: VY012451.D
Acq: 02 Feb 2023 11:58




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 43 Resp: 12843
Ion Ratio Lower Upper
43 100
61 32.8 24.4 36.6
87 16.9 11.4 17.2




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 02/03/2023
Supervised By :Mahesh Dadoda 02/03/2023



#44
Trichloroethene
Concen: 52.379 ug/l
RT: 8.941 min Scan# 1168
Delta R.T. -0.000 min
Lab File: VY012451.D
Acq: 02 Feb 2023 11:58

Tgt Ion:130 Resp: 122924
Ion Ratio Lower Upper
130 100
95 92.1 0.0 190.6





#45

1,2-Dichloropropane

Concen: 57.838 ug/l

RT: 9.221 min Scan# 1214

Delta R.T. 0.005 min

Lab File: VY012451.D

Acq: 02 Feb 2023 11:58

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion: 63 Resp: 9475

Ion Ratio Lower Upper

63 100





65 32.0 24.9 37.3

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 02/03/2023

Supervised By :Mahesh Dadoda 02/03/2023



#46

Dibromomethane

Concen: 52.927 ug/l

RT: 9.307 min Scan# 1228

Delta R.T. -0.000 min

Lab File: VY012451.D

Acq: 02 Feb 2023 11:58




Tgt Ion: 93 Resp: 58050


Ion Ratio Lower Upper

93 100

95 83.6 66.1 99.1

174 113.8 84.7 127.1









#47
Bromodichloromethane
Concen: 54.050 ug/l
RT: 9.496 min Scan# 1259
Delta R.T. -0.000 min
Lab File: VY012451.D
Acq: 02 Feb 2023 11:58

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion: 83 Resp: 14922
Ion Ratio Lower Upper
83 100
85 64.2 52.0 78.0
127 9.3 7.3 10.9




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 02/03/2023
Supervised By :Mahesh Dadoda 02/03/2023



#48
Methyl methacrylate
Concen: 57.397 ug/l
RT: 9.295 min Scan# 1226
Delta R.T. -0.000 min
Lab File: VY012451.D
Acq: 02 Feb 2023 11:58

Tgt Ion: 41 Resp: 59006
Ion Ratio Lower Upper
41 100
69 103.5 74.2 111.2
39 58.5 46.3 69.5





#49

1,4-Dioxane

Concen: 1115.943 ug/l

RT: 9.301 min Scan# 1227

Delta R.T. 0.006 min

Lab File: VY012451.D

Acq: 02 Feb 2023 11:58

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion: 88 Resp: 16678

Ion Ratio Lower Upper

88 100

43 25.7 26.2 39.2





58 64.7 54.3 81.5

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 02/03/2023

Supervised By :Mahesh Dadoda 02/03/2023



#50

Toluene-d8

Concen: 45.647 ug/l

RT: 10.185 min Scan# 1372

Delta R.T. 0.006 min

Lab File: VY012451.D




Acq: 02 Feb 2023 11:58


Tgt Ion: 98 Resp: 255249

Ion Ratio Lower Upper

98 100

100 65.2 51.9 77.9









#51
4-Methyl-2-Pentanone
Concen: 290.672 ug/l
RT: 10.069 min Scan# 1353
Delta R.T. -0.000 min
Lab File: VY012451.D
Acq: 02 Feb 2023 11:58

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion: 43 Resp: 34582
Ion Ratio Lower Upper
43 100
58 44.4 32.9 49.3




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 02/03/2023
Supervised By :Mahesh Dadoda 02/03/2023



#52
Toluene
Concen: 52.945 ug/l
RT: 10.246 min Scan# 1382
Delta R.T. -0.000 min
Lab File: VY012451.D
Acq: 02 Feb 2023 11:58


Tgt Ion: 92 Resp: 276786
Ion Ratio Lower Upper
92 100
91 170.1 137.4 206.0





#53
 t-1,3-Dichloropropene
 Concen: 53.291 ug/l
 RT: 10.465 min Scan# 1418
 Delta R.T. 0.000 min
 Lab File: VY012451.D
 Acq: 02 Feb 2023 11:58




Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050



Tgt Ion: 75 Resp: 153900
 Ion Ratio Lower Upper
 75 100
 77 32.6 25.8 38.8




Manual Integrations
 APPROVED


Reviewed By :Krupa Patel 02/03/2023
 Supervised By :Mahesh Dadoda 02/03/2023



#54
 cis-1,3-Dichloropropene
 Concen: 54.575 ug/l
 RT: 9.929 min Scan# 1330
 Delta R.T. -0.000 min
 Lab File: VY012451.D
 Acq: 02 Feb 2023 11:58


Tgt Ion: 75 Resp: 169734
 Ion Ratio Lower Upper
 75 100
 77 31.6 25.9 38.9
 39 42.8 34.6 52.0





#55
1,1,2-Trichloroethane
Concen: 54.932 ug/l
RT: 10.648 min Scan# 1447
Delta R.T. 0.006 min
Lab File: VY012451.D
Acq: 02 Feb 2023 11:58




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion	Ratio	Lower	Upper
97	100		
83	80.3	67.7	101.5
85	52.8	43.8	65.8
99	61.2	49.0	73.4

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 02/03/2023
Supervised By :Mahesh Dadoda 02/03/2023




#56
Ethyl methacrylate
Concen: 57.146 ug/l
RT: 10.514 min Scan# 1426
Delta R.T. 0.006 min
Lab File: VY012451.D
Acq: 02 Feb 2023 11:58



Tgt Ion	Ratio	Lower	Upper
69	100		
41	54.8	48.9	73.3
39	31.4	25.9	38.9





#57

1,3-Dichloropropane

Concen: 55.876 ug/l

RT: 10.794 min Scan# 1471

Delta R.T. 0.005 min

Lab File: VY012451.D

Acq: 02 Feb 2023 11:58

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion: 76 Resp: 14220

Ion Ratio Lower Upper

76 100





78 32.5 25.8 38.8

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 02/03/2023

Supervised By :Mahesh Dadoda 02/03/2023



#58

2-Chloroethyl Vinyl ether

Concen: 259.699 ug/l

RT: 9.782 min Scan# 1306

Delta R.T. -0.001 min

Lab File: VY012451.D




Acq: 02 Feb 2023 11:58


Tgt Ion: 63 Resp: 175675

Ion Ratio Lower Upper

63 100


106 33.3 23.2 34.8





#59
2-Hexanone
Concen: 274.463 ug/l
RT: 10.831 min Scan# 1478
Delta R.T. -0.000 min
Lab File: VY012451.D
Acq: 02 Feb 2023 11:58




Instrument :
MSVOA_Y
Client Sample Id :
VSTDCCC050




Tgt Ion: 43 Resp: 248609
Ion Ratio Lower Upper
43 100
58 60.9 28.4 85.2

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 02/03/2023
Supervised By :Mahesh Dadoda 02/03/2023




#60
Dibromochloromethane
Concen: 54.181 ug/l
RT: 10.989 min Scan# 1504
Delta R.T. 0.005 min
Lab File: VY012451.D
Acq: 02 Feb 2023 11:58



Tgt Ion:129 Resp: 111981
Ion Ratio Lower Upper
129 100
127 77.1 38.9 116.6









#61
1,2-Dibromoethane
Concen: 54.176 ug/l
RT: 11.093 min Scan# 1520
Delta R.T. 0.006 min
Lab File: VY012451.D
Acq: 02 Feb 2023 11:58

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion:107 Resp: 8146
Ion Ratio Lower Upper
107 100
109 94.2 75.8 113.6




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 02/03/2023
Supervised By :Mahesh Dadoda 02/03/2023



#62
4-Bromofluorobenzene
Concen: 45.360 ug/l
RT: 12.483 min Scan# 1749
Delta R.T. 0.000 min
Lab File: VY012451.D
Acq: 02 Feb 2023 11:58

Tgt Ion: 95 Resp: 102352
Ion Ratio Lower Upper
95 100
174 90.1 0.0 172.0
176 87.1 0.0 167.6





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.495 min Scan# 11

Delta R.T. 0.005 min

Lab File: VY012451.D

Acq: 02 Feb 2023 11:58

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion:117 Resp: 272875





Ion	Ratio	Lower	Upper
117	100		
82	52.5	44.8	67.2
119	32.1	25.4	38.2

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 02/03/2023

Supervised By :Mahesh Dadoda 02/03/2023



#64

Tetrachloroethene

Concen: 54.550 ug/l

RT: 10.721 min Scan# 1460




Delta R.T. 0.000 min


Lab File: VY012451.D

Acq: 02 Feb 2023 11:58

Tgt Ion:164 Resp: 135826

Ion	Ratio	Lower	Upper
164	100		
166	125.7	102.3	153.5
129	93.2	77.2	115.8
131	88.5	72.5	108.7









#65
Chlorobenzene
Concen: 53.506 ug/l
RT: 11.520 min Scan# 1150
Delta R.T. 0.006 min
Lab File: VY012451.D
Acq: 02 Feb 2023 11:58

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion:112 Resp: 29873
Ion Ratio Lower Upper
112 100
114 32.2 25.3 37.9




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 02/03/2023
Supervised By :Mahesh Dadoda 02/03/2023



#66
1,1,1,2-Tetrachloroethane
Concen: 54.492 ug/l
RT: 11.593 min Scan# 1603
Delta R.T. 0.006 min
Lab File: VY012451.D
Acq: 02 Feb 2023 11:58


Tgt Ion:131 Resp: 114836
Ion Ratio Lower Upper
131 100
133 95.1 47.3 142.0
119 61.4 32.1 96.3





#67
Ethyl Benzene
Concen: 53.637 ug/l
RT: 11.593 min Scan# 1603
Delta R.T. -0.000 min
Lab File: VY012451.D
Acq: 02 Feb 2023 11:58




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 91 Resp: 535197
Ion Ratio Lower Upper
91 100
106 31.4 25.0 37.4




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 02/03/2023
Supervised By :Mahesh Dadoda 02/03/2023



#68
m/p-Xylenes
Concen: 106.625 ug/l
RT: 11.703 min Scan# 1621
Delta R.T. -0.000 min
Lab File: VY012451.D
Acq: 02 Feb 2023 11:58


Tgt Ion:106 Resp: 420660
Ion Ratio Lower Upper
106 100
91 200.3 162.5 243.7





#69
o-Xylene
Concen: 53.607 ug/l
RT: 12.032 min Scan# 1675
Delta R.T. 0.006 min
Lab File: VY012451.D
Acq: 02 Feb 2023 11:58




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion:106 Resp: 20200
Ion Ratio Lower Upper
106 100
91 212.1 107.5 322.4

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 02/03/2023
Supervised By :Mahesh Dadoda 02/03/2023




#70
Styrene
Concen: 54.373 ug/l
RT: 12.044 min Scan# 1677
Delta R.T. 0.000 min
Lab File: VY012451.D
Acq: 02 Feb 2023 11:58



Tgt Ion:104 Resp: 342341
Ion Ratio Lower Upper
104 100
78 48.5 39.4 59.0
103 54.1 43.0 64.4





#71

Bromoform

Concen: 53.604 ug/l

RT: 12.209 min Scan# 1704

Delta R.T. -0.000 min

Lab File: VY012451.D

Acq: 02 Feb 2023 11:58

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050





Tgt Ion:	173	Resp:	7217
Ion Ratio	Lower	Upper	
173	100		
175	49.3	23.8	71.5
254	0.1	0.1	0.1

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 02/03/2023

Supervised By :Mahesh Dadoda 02/03/2023



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l




RT: 13.428 min Scan# 1904


Delta R.T. 0.006 min

Lab File: VY012451.D

Acq: 02 Feb 2023 11:58

Tgt Ion:	152	Resp:	134846
Ion Ratio	Lower	Upper	
152	100		
115	57.8	29.1	87.3
150	175.7	0.0	342.6





#73

Isopropylbenzene

Concen: 55.799 ug/l

RT: 12.331 min Scan# 11

Delta R.T. -0.000 min

Lab File: VY012451.D

Acq: 02 Feb 2023 11:58

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion:105 Resp: 54256

Ion Ratio Lower Upper

105 100





120 27.0 13.2 39.6

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 02/03/2023

Supervised By :Mahesh Dadoda 02/03/2023



#74

N-ethyl acetate

Concen: 59.024 ug/l

RT: 12.148 min Scan# 1694

Delta R.T. 0.006 min

Lab File: VY012451.D

Acq: 02 Feb 2023 11:58

Tgt Ion: 43 Resp: 112506




Ion Ratio Lower Upper


43 100

70 50.5 35.8 53.8

55 26.6 21.0 31.4

61 26.9 20.0 30.0









#75
1,1,2,2-Tetrachloroethane
Concen: 58.850 ug/l
RT: 12.587 min Scan# 1766
Delta R.T. 0.006 min
Lab File: VY012451.D
Acq: 02 Feb 2023 11:58

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion: 83 Resp: 9013
Ion Ratio Lower Upper
83 100
131 12.2 5.6 16.8
85 64.5 32.3 96.8




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 02/03/2023
Supervised By :Mahesh Dadoda 02/03/2023



#76
1,2,3-Trichloropropane
Concen: 64.567 ug/l m
RT: 12.635 min Scan# 1774
Delta R.T. 0.005 min
Lab File: VY012451.D
Acq: 02 Feb 2023 11:58

Tgt Ion: 75 Resp: 80060
Ion Ratio Lower Upper
75 100
77 0.0 0.0 0.0





#77

Bromobenzene

Concen: 54.939 ug/l

RT: 12.611 min Scan# 1770

Delta R.T. 0.006 min

Lab File: VY012451.D

Acq: 02 Feb 2023 11:58

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050





Tgt Ion:	156	Resp:	128540
Ion Ratio	Lower	Upper	
156	100		
77	165.0	90.0	270.0
158	96.3	48.4	145.2

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 02/03/2023

Supervised By :Mahesh Dadoda 02/03/2023



#78

n-propylbenzene

Concen: 55.752 ug/l




RT: 12.672 min Scan# 1780


Delta R.T. 0.000 min

Lab File: VY012451.D

Acq: 02 Feb 2023 11:58


Tgt Ion:	91	Resp:	634373
Ion Ratio	Lower	Upper	
91	100		
120	24.0	11.8	35.4





#79
2-Chlorotoluene
Concen: 55.160 ug/l
RT: 12.757 min Scan# 1794
Delta R.T. -0.001 min
Lab File: VY012451.D
Acq: 02 Feb 2023 11:58




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion: 91 Resp: 35486
Ion Ratio Lower Upper
91 100
126 35.3 17.2 51.5

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 02/03/2023
Supervised By :Mahesh Dadoda 02/03/2023




#80
1,3,5-Trimethylbenzene
Concen: 54.560 ug/l
RT: 12.812 min Scan# 1803
Delta R.T. -0.001 min
Lab File: VY012451.D
Acq: 02 Feb 2023 11:58



Tgt Ion:105 Resp: 461193
Ion Ratio Lower Upper
105 100
120 49.3 24.2 72.6









#81
trans-1,4-Dichloro-2-butene
Concen: 56.928 ug/l
RT: 12.379 min Scan# 1732
Delta R.T. 0.005 min
Lab File: VY012451.D
Acq: 02 Feb 2023 11:58

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion: 75 Resp: 34259
Ion Ratio Lower Upper
75 100
53 89.2 72.5 108.7
89 49.3 37.0 55.4




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 02/03/2023
Supervised By :Mahesh Dadoda 02/03/2023



#82
4-Chlorotoluene
Concen: 53.797 ug/l
RT: 12.855 min Scan# 1810
Delta R.T. -0.000 min
Lab File: VY012451.D
Acq: 02 Feb 2023 11:58

Tgt Ion: 91 Resp: 363543
Ion Ratio Lower Upper
91 100
126 35.6 17.1 51.2





#83

tert-Butylbenzene

Concen: 55.440 ug/l

RT: 13.080 min Scan# 1846

Delta R.T. 0.005 min

Lab File: VY012451.D

Acq: 02 Feb 2023 11:58

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion:119 Resp: 40844

Ion Ratio Lower Upper

119 100

91 64.2 32.3 96.8





134 25.2 13.0 39.0

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 02/03/2023

Supervised By :Mahesh Dadoda 02/03/2023



#84

1,2,4-Trimethylbenzene

Concen: 54.301 ug/l

RT: 13.123 min Scan# 1854

Delta R.T. 0.006 min

Lab File: VY012451.D




Acq: 02 Feb 2023 11:58


Tgt Ion:105 Resp: 452658

Ion Ratio Lower Upper

105 100

120 45.4 22.9 68.7









#85
 sec-Butylbenzene
 Concen: 55.765 ug/l
 RT: 13.257 min Scan# 1875
 Delta R.T. 0.006 min
 Lab File: VY012451.D
 Acq: 02 Feb 2023 11:58

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050

Tgt Ion:105 Resp: 584929
 Ion Ratio Lower Upper
 105 100
 134 21.0 10.1 30.3




Manual Integrations
 APPROVED


Reviewed By :Krupa Patel 02/03/2023
 Supervised By :Mahesh Dadoda 02/03/2023



#86
 p-Isopropyltoluene
 Concen: 54.244 ug/l
 RT: 13.373 min Scan# 1895
 Delta R.T. 0.006 min
 Lab File: VY012451.D
 Acq: 02 Feb 2023 11:58


Tgt Ion:119 Resp: 497115
 Ion Ratio Lower Upper
 119 100
 134 26.5 13.0 39.0
 91 23.4 11.9 35.9





#87
1,3-Dichlorobenzene
Concen: 53.570 ug/l
RT: 13.367 min Scan# 1894
Delta R.T. 0.000 min
Lab File: VY012451.D
Acq: 02 Feb 2023 11:58




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion:146 Resp: 249292
Ion Ratio Lower Upper
146 100
111 39.1 19.7 59.0
148 63.6 32.0 96.0

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 02/03/2023
Supervised By :Mahesh Dadoda 02/03/2023




#88
1,4-Dichlorobenzene
Concen: 53.295 ug/l
RT: 13.446 min Scan# 1907
Delta R.T. -0.001 min
Lab File: VY012451.D
Acq: 02 Feb 2023 11:58



Tgt Ion:146 Resp: 247339
Ion Ratio Lower Upper
146 100
111 37.8 19.6 58.7
148 63.7 31.9 95.7





#89

n-Butylbenzene

Concen: 55.529 ug/l

RT: 13.696 min Scan# 1948

Delta R.T. -0.001 min

Lab File: VY012451.D

Acq: 02 Feb 2023 11:58

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050





Tgt Ion:	91	Resp:	44101
Ion Ratio	Lower	Upper	
91	100		
92	53.0	26.8	80.4
134	27.8	13.3	39.8

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 02/03/2023

Supervised By :Mahesh Dadoda 02/03/2023



#90

Hexachloroethane

Concen: 54.938 ug/l




RT: 13.964 min Scan# 1992


Delta R.T. 0.005 min

Lab File: VY012451.D

Acq: 02 Feb 2023 11:58

Tgt Ion:	117	Resp:	90891
Ion Ratio	Lower	Upper	
117	100		
201	85.2	40.3	120.8









#91
1,2-Dichlorobenzene
Concen: 54.022 ug/l
RT: 13.745 min Scan# 1955
Delta R.T. 0.006 min
Lab File: VY012451.D
Acq: 02 Feb 2023 11:58

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion:146 Resp: 224314
Ion Ratio Lower Upper
146 100
111 40.1 20.3 60.8
148 63.7 31.8 95.4




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 02/03/2023
Supervised By :Mahesh Dadoda 02/03/2023



#92
1,2-Dibromo-3-Chloropropane
Concen: 53.289 ug/l
RT: 14.361 min Scan# 2057
Delta R.T. 0.006 min
Lab File: VY012451.D
Acq: 02 Feb 2023 11:58

Tgt Ion: 75 Resp: 16867
Ion Ratio Lower Upper
75 100
155 92.0 42.3 126.8
157 117.9 54.4 163.3





#93

1,2,4-Trichlorobenzene

Concen: 51.251 ug/l

RT: 15.007 min Scan# 2163

Delta R.T. -0.000 min

Lab File: VY012451.D

Acq: 02 Feb 2023 11:58

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion:180 Resp: 13635





Ion	Ratio	Lower	Upper
180	100		
182	96.0	48.1	144.3
145	30.9	15.4	46.4

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 02/03/2023

Supervised By :Mahesh Dadoda 02/03/2023



#94

Hexachlorobutadiene

Concen: 50.895 ug/l

RT: 15.111 min Scan# 2180




Delta R.T. -0.000 min


Lab File: VY012451.D

Acq: 02 Feb 2023 11:58

Tgt Ion:225 Resp: 81527

Ion	Ratio	Lower	Upper
225	100		
223	63.9	31.4	94.2
227	64.9	32.6	97.8





#95

Naphthalene

Concen: 53.313 ug/l

RT: 15.239 min Scan# 2201

Delta R.T. 0.006 min

Lab File: VY012451.D

Acq: 02 Feb 2023 11:58

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion:128 Resp: 27996

Ion Ratio Lower Upper

128 100

127 12.9 10.2 15.2





129 11.1 8.8 13.2

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 02/03/2023

Supervised By :Mahesh Dadoda 02/03/2023



#96

1,2,3-Trichlorobenzene

Concen: 51.643 ug/l

RT: 15.428 min Scan# 2232

Delta R.T. 0.006 min

Lab File: VY012451.D

Acq: 02 Feb 2023 11:58

Tgt Ion:180 Resp: 118236

Ion Ratio Lower Upper

180 100

182 95.5 47.9 143.7

145 32.4 16.1 48.2

