

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020224\
 Data File : VY017180.D
 Acq On : 02 Feb 2024 13:40
 Operator : SY/MD
 Sample : VIBLK
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VIBLK

Quant Time: Feb 02 22:29:24 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020124S.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 02 00:56:04 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	93	50.000	ug/l	#-0.01
34) 1,4-Difluorobenzene	8.703	114	720	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.502	117	1148	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	1087	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	164	197.399	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	394.800%#
35) Dibromofluoromethane	7.728	113	44	11.318	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	22.640%#
50) Toluene-d8	10.191	98	367	25.929	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	51.860%#
62) 4-Bromofluorobenzene	12.489	95	416	89.720	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	179.440%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020224\
Data File : VY017180.D
Acq On : 02 Feb 2024 13:40
Operator : SY/MD
Sample : VIBLK
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VIBLK

Quant Time: Feb 02 22:29:24 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020124S.M
Quant Title : SW846 8260
QLast Update : Fri Feb 02 00:56:04 2024
Response via : Initial Calibration

