

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020224\
 Data File : VY017187.D
 Acq On : 02 Feb 2024 16:24
 Operator : SY/MD
 Sample : P1343-11
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BP-VPB-195-GW-660-662

Quant Time: Feb 02 22:31:58 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020124S.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 02 00:56:04 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	81828	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	176882	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	178370	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	69543	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.155	65	43991	60.179	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	120.360%
35) Dibromofluoromethane	7.728	113	43772	45.832	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	91.660%
50) Toluene-d8	10.191	98	155329	44.670	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	89.340%
62) 4-Bromofluorobenzene	12.489	95	54332	47.698	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	95.400%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020224\
Data File : VY017187.D
Acq On : 02 Feb 2024 16:24
Operator : SY/MD
Sample : P1343-11
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BP-VPB-195-GW-660-662

Quant Time: Feb 02 22:31:58 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020124S.M
Quant Title : SW846 8260
QLast Update : Fri Feb 02 00:56:04 2024
Response via : Initial Calibration

