

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020224\
 Data File : VY017192.D
 Acq On : 02 Feb 2024 18:22
 Operator : SY/MD
 Sample : P1297-01
 Misc : 4.99g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TR-05-013024

Quant Time: Feb 02 22:33:36 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020124S.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 02 00:56:04 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	62552	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	124975	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	118684	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	44052	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	27069	48.441	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	96.880%
35) Dibromofluoromethane	7.722	113	29441	43.630	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	87.260%
50) Toluene-d8	10.185	98	109162	44.432	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	88.860%
62) 4-Bromofluorobenzene	12.489	95	32197	40.005	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	80.020%
Target Compounds						Qvalue
16) Acetone	3.948	43	3957	24.093	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020224\
Data File : VY017192.D
Acq On : 02 Feb 2024 18:22
Operator : SY/MD
Sample : P1297-01
Misc : 4.99g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TR-05-013024

Quant Time: Feb 02 22:33:36 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020124S.M
Quant Title : SW846 8260
QLast Update : Fri Feb 02 00:56:04 2024
Response via : Initial Calibration

