

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020224\
 Data File : VY017202.D
 Acq On : 02 Feb 2024 23:46
 Operator : SY/MD
 Sample : P1313-02
 Misc : 3.71g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 COMP-2

Quant Time: Feb 03 01:43:24 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020124S.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 02 00:56:04 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.807	168	82132	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	8.703	114	174637	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.508	117	172081	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.440	152	64447	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.161	65	39742	54.165	ug/l	0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	=	108.340%
35) Dibromofluoromethane	7.734	113	41727	44.253	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	88.500%
50) Toluene-d8	10.197	98	153711	44.773	ug/l	0.01
Spiked Amount	50.000	Range	58 - 134	Recovery	=	89.540%
62) 4-Bromofluorobenzene	12.495	95	49204	43.751	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	87.500%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020224\
Data File : VY017202.D
Acq On : 02 Feb 2024 23:46
Operator : SY/MD
Sample : P1313-02
Misc : 3.71g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
COMP-2

Quant Time: Feb 03 01:43:24 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020124S.M
Quant Title : SW846 8260
QLast Update : Fri Feb 02 00:56:04 2024
Response via : Initial Calibration

