

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020224\
 Data File : VY017204.D
 Acq On : 03 Feb 2024 00:33
 Operator : SY/MD
 Sample : P1313-01
 Misc : 5.01g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 COMP-1

Quant Time: Feb 03 01:44:07 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020124S.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 02 00:56:04 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.813	168	81345	50.000	ug/l	0.02
34) 1,4-Difluorobenzene	8.709	114	174888	50.000	ug/l	0.01
63) Chlorobenzene-d5	11.508	117	176582	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.440	152	67903	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.167	65	40330	55.499	ug/l	0.02
Spiked Amount	50.000	Range	50 - 163	Recovery	=	111.000%
35) Dibromofluoromethane	7.740	113	42892	45.423	ug/l	0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	90.840%
50) Toluene-d8	10.197	98	153837	44.745	ug/l	0.01
Spiked Amount	50.000	Range	58 - 134	Recovery	=	89.500%
62) 4-Bromofluorobenzene	12.501	95	50550	44.884	ug/l	0.01
Spiked Amount	50.000	Range	30 - 143	Recovery	=	89.760%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020224\
Data File : VY017204.D
Acq On : 03 Feb 2024 00:33
Operator : SY/MD
Sample : P1313-01
Misc : 5.01g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
COMP-1

Quant Time: Feb 03 01:44:07 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020124S.M
Quant Title : SW846 8260
QLast Update : Fri Feb 02 00:56:04 2024
Response via : Initial Calibration

