

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020224\
 Data File : VY017213.D
 Acq On : 03 Feb 2024 04:05
 Operator : SY/MD
 Sample : P1345-03
 Misc : 5.17g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TT-278-IDWSO-20240201-3

Quant Time: Feb 03 04:31:21 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020124S.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 02 00:56:04 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.819	168	81810	50.000	ug/l	0.02
34) 1,4-Difluorobenzene	8.715	114	177274	50.000	ug/l	0.02
63) Chlorobenzene-d5	11.508	117	186416	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.446	152	75965	50.000	ug/l	0.01
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.167	65	41312	56.527	ug/l	0.02
Spiked Amount	50.000	Range	50 - 163	Recovery	=	113.060%
35) Dibromofluoromethane	7.740	113	43112	45.041	ug/l	0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	90.080%
50) Toluene-d8	10.203	98	158358	45.440	ug/l	0.02
Spiked Amount	50.000	Range	58 - 134	Recovery	=	90.880%
62) 4-Bromofluorobenzene	12.501	95	57912	50.728	ug/l	0.01
Spiked Amount	50.000	Range	30 - 143	Recovery	=	101.460%
Target Compounds						
20) Methylene Chloride	4.747	84	10088	8.563	ug/l	Qvalue 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020224\
Data File : VY017213.D
Acq On : 03 Feb 2024 04:05
Operator : SY/MD
Sample : P1345-03
Misc : 5.17g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TT-278-IDWSO-20240201-3

Quant Time: Feb 03 04:31:21 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020124S.M
Quant Title : SW846 8260
QLast Update : Fri Feb 02 00:56:04 2024
Response via : Initial Calibration

