

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020224\
 Data File : VY017197.D
 Acq On : 02 Feb 2024 21:04
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050

Quant Time: Feb 02 22:35:17 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020124S.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 02 00:56:04 2024
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Semsettin Yesilyurt 02/05/2024
 Supervised By :Mahesh Dadoda 02/05/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	114068	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	198926	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.502	117	173654	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	79382	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	53007	52.018	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery = 104.040%			
35) Dibromofluoromethane	7.722	113	55060	51.263	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery = 102.520%			
50) Toluene-d8	10.185	98	207251	52.997	ug/l	0.00
Spiked Amount 50.000	Range 58 - 134		Recovery = 106.000%			
62) 4-Bromofluorobenzene	12.489	95	67005	52.305	ug/l	0.00
Spiked Amount 50.000	Range 30 - 143		Recovery = 104.600%			
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.906	85	39804	46.031	ug/l	96
3) Chloromethane	2.113	50	80347	51.723	ug/l	95
4) Vinyl Chloride	2.253	62	92731	48.985	ug/l	99
5) Bromomethane	2.650	94	67481	51.425	ug/l	98
6) Chloroethane	2.796	64	62923	51.480	ug/l	100
7) Trichlorofluoromethane	3.125	101	94677	48.637	ug/l	99
8) Diethyl Ether	3.534	74	33545	54.785	ug/l	99
9) 1,1,2-Trichlorotrifluo...	3.906	101	58320	49.424	ug/l	99
10) Methyl Iodide	4.095	142	66828	53.870	ug/l	97
11) Tert butyl alcohol	4.960	59	20721	240.999	ug/l	98
12) 1,1-Dichloroethene	3.875	96	56590	51.181	ug/l	94
13) Acrolein	3.735	56	19164	225.837	ug/l	99
14) Allyl chloride	4.485	41	90427	54.098	ug/l	99
15) Acrylonitrile	5.167	53	73982	268.358	ug/l	100
16) Acetone	3.954	43	62852	209.855	ug/l	98
17) Carbon Disulfide	4.198	76	183229	51.914	ug/l	100
18) Methyl Acetate	4.485	43	41502	62.116	ug/l	100
19) Methyl tert-butyl Ether	5.228	73	143901	54.039	ug/l	99
20) Methylene Chloride	4.722	84	69877	57.490	ug/l	99
21) trans-1,2-Dichloroethene	5.228	96	68829	54.720	ug/l	99
22) Diisopropyl ether	6.125	45	204825	55.799	ug/l	97
23) Vinyl Acetate	6.070	43	584856	262.954	ug/l	100
24) 1,1-Dichloroethane	6.027	63	121913	54.215	ug/l	99
25) 2-Butanone	6.990	43	95633	240.193	ug/l	97
26) 2,2-Dichloropropane	6.990	77	93683	50.087	ug/l	99
27) cis-1,2-Dichloroethene	6.990	96	76843	54.622	ug/l	99
28) Bromochloromethane	7.344	49	48145	55.917	ug/l	99
29) Tetrahydrofuran	7.356	42	60120	267.302	ug/l	100
30) Chloroform	7.515	83	123845	54.748	ug/l	98
31) Cyclohexane	7.795	56	96826	47.142	ug/l	99
32) 1,1,1-Trichloroethane	7.710	97	102638	52.749	ug/l	99
36) 1,1-Dichloropropene	7.923	75	94710	51.120	ug/l	99
37) Ethyl Acetate	7.082	43	41984	48.858	ug/l	98
38) Carbon Tetrachloride	7.911	117	89424	50.411	ug/l	99
39) Methylcyclohexane	9.191	83	105853	48.905	ug/l	98
40) Benzene	8.167	78	281553	53.216	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020224\
 Data File : VY017197.D
 Acq On : 02 Feb 2024 21:04
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 02/05/2024
 Supervised By :Mahesh Dadoda 02/05/2024

Quant Time: Feb 02 22:35:17 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020124S.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 02 00:56:04 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.320	41	24498	60.172	ug/l	99
42) 1,2-Dichloroethane	8.246	62	72676	51.881	ug/l	100
43) Isopropyl Acetate	8.283	43	78786	51.794	ug/l	100
44) Trichloroethene	8.947	130	71934	53.561	ug/l	94
45) 1,2-Dichloropropane	9.222	63	68352	52.596	ug/l	99
46) Dibromomethane	9.313	93	36816	52.445	ug/l	99
47) Bromodichloromethane	9.502	83	93285	53.098	ug/l	96
48) Methyl methacrylate	9.301	41	36906	51.725	ug/l	96
49) 1,4-Dioxane	9.307	88	8282	1083.912	ug/l	97
51) 4-Methyl-2-Pentanone	10.081	43	216502	265.514	ug/l	99
52) Toluene	10.252	92	172507	54.017	ug/l	99
53) t-1,3-Dichloropropene	10.471	75	86909	52.666	ug/l	98
54) cis-1,3-Dichloropropene	9.935	75	104771	53.006	ug/l	99
55) 1,1,2-Trichloroethane	10.654	97	51060	53.460	ug/l	98
56) Ethyl methacrylate	10.520	69	64213	49.971	ug/l	99
57) 1,3-Dichloropropane	10.801	76	88372	53.795	ug/l	99
58) 2-Chloroethyl Vinyl ether	9.789	63	170269	286.159	ug/l	99
59) 2-Hexanone	10.837	43	147090	249.464	ug/l	100
60) Dibromochloromethane	10.996	129	59271	52.898	ug/l	100
61) 1,2-Dibromoethane	11.099	107	47005	52.917	ug/l	99
64) Tetrachloroethene	10.727	164	60868	53.971	ug/l	97
65) Chlorobenzene	11.526	112	176939	52.819	ug/l	99
66) 1,1,1,2-Tetrachloroethane	11.599	131	65807	54.055	ug/l	99
67) Ethyl Benzene	11.599	91	326277	53.892	ug/l	100
68) m/p-Xylenes	11.709	106	250324	108.951	ug/l	100
69) o-Xylene	12.038	106	116800	55.203	ug/l	100
70) Styrene	12.050	104	196595	56.394	ug/l	99
71) Bromoform	12.215	173	33962	53.269	ug/l #	99
73) Isopropylbenzene	12.337	105	304710	52.251	ug/l	100
74) N-amyl acetate	12.154	43	67647	51.625	ug/l	99
75) 1,1,2,2-Tetrachloroethane	12.593	83	57165	49.839	ug/l	99
76) 1,2,3-Trichloropropane	12.642	75	37703m	51.459	ug/l	
77) Bromobenzene	12.617	156	68050	52.607	ug/l	96
78) n-propylbenzene	12.678	91	375160	52.379	ug/l	100
79) 2-Chlorotoluene	12.764	91	207449	52.683	ug/l	99
80) 1,3,5-Trimethylbenzene	12.825	105	247405	53.343	ug/l	98
81) trans-1,4-Dichloro-2-b...	12.386	75	17203	48.787	ug/l	99
82) 4-Chlorotoluene	12.867	91	211449	52.176	ug/l	99
83) tert-Butylbenzene	13.087	119	215582	54.226	ug/l	99
84) 1,2,4-Trimethylbenzene	13.129	105	245271	53.576	ug/l	99
85) sec-Butylbenzene	13.264	105	320805	51.629	ug/l	99
86) p-Isopropyltoluene	13.379	119	260738	52.511	ug/l	99
87) 1,3-Dichlorobenzene	13.373	146	134770	52.226	ug/l	99
88) 1,4-Dichlorobenzene	13.459	146	132598	52.890	ug/l	100
89) n-Butylbenzene	13.703	91	247701	52.468	ug/l	99
90) Hexachloroethane	13.971	117	52426	51.431	ug/l	98
91) 1,2-Dichlorobenzene	13.751	146	116841	53.187	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.367	75	7444	49.825	ug/l	97
93) 1,2,4-Trichlorobenzene	15.019	180	61935	56.760	ug/l	99
94) Hexachlorobutadiene	15.123	225	33448	52.046	ug/l	99
95) Naphthalene	15.245	128	116843	52.674	ug/l	100
96) 1,2,3-Trichlorobenzene	15.434	180	52849	56.947	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020224\
Data File : VY017197.D
Acq On : 02 Feb 2024 21:04
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/05/2024
Supervised By :Mahesh Dadoda 02/05/2024

Quant Time: Feb 02 22:35:17 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020124S.M
Quant Title : SW846 8260
QLast Update : Fri Feb 02 00:56:04 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020224\
 Data File : VY017197.D
 Acq On : 02 Feb 2024 21:04
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 28 Sample Multiplier: 1


Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050

Quant Time: Feb 02 22:35:17 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020124S.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 02 00:56:04 2024
 Response via : Initial Calibration

Manual Integrations APPROVED


Reviewed By :Semsettin Yesilyurt 02/05/2024
 Supervised By :Mahesh Dadoda 02/05/2024





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.801 min Scan# 981
Delta R.T. 0.006 min
Lab File: VY017197.D
Acq: 02 Feb 2024 21:04




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion:168 Resp: 114068
Ion Ratio Lower Upper
168 100
99 51.9 44.2 66.4




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 02/05/2024
Supervised By :Mahesh Dadoda 02/05/2024



#2
Dichlorodifluoromethane
Concen: 46.031 ug/l
RT: 1.906 min Scan# 14
Delta R.T. 0.000 min
Lab File: VY017197.D
Acq: 02 Feb 2024 21:04

Tgt Ion: 85 Resp: 39804
Ion Ratio Lower Upper
85 100
87 32.1 17.1 51.3





#3

Chloromethane

Concen: 51.723 ug/l

RT: 2.113 min Scan# 41

Delta R.T. -0.006 min

Lab File: VY017197.D

Acq: 02 Feb 2024 21:04

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion: 50 Resp: 8034

Ion Ratio Lower Upper

50 100





52 30.8 26.7 40.1

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 02/05/2024

Supervised By :Mahesh Dadoda 02/05/2024



#4

Vinyl Chloride

Concen: 48.985 ug/l

RT: 2.253 min Scan# 71

Delta R.T. 0.000 min

Lab File: VY017197.D




Acq: 02 Feb 2024 21:04


Tgt Ion: 62 Resp: 92731

Ion Ratio Lower Upper

62 100

64 31.7 25.7 38.5





#5

Bromomethane

Concen: 51.425 ug/l

RT: 2.650 min Scan# 11

Delta R.T. -0.006 min

Lab File: VY017197.D

Acq: 02 Feb 2024 21:04

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion: 94 Resp: 6748

Ion Ratio Lower Upper

94 100





96 94.2 77.3 115.9

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 02/05/2024

Supervised By :Mahesh Dadoda 02/05/2024



#6

Chloroethane

Concen: 51.480 ug/l

RT: 2.796 min Scan# 160

Delta R.T. 0.000 min

Lab File: VY017197.D




Acq: 02 Feb 2024 21:04


Tgt Ion: 64 Resp: 62923

Ion Ratio Lower Upper

64 100


66 31.2 25.1 37.7





#7
 Trichlorofluoromethane
 Concen: 48.637 ug/l
 RT: 3.125 min Scan# 215
 Delta R.T. -0.006 min
 Lab File: VY017197.D
 Acq: 02 Feb 2024 21:04




Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050



Tgt Ion:101 Resp: 9467
 Ion Ratio Lower Upper
 101 100
 103 65.8 51.9 77.9




Manual Integrations
 APPROVED


Reviewed By :Semsettin Yesilyurt 02/05/2024
 Supervised By :Mahesh Dadoda 02/05/2024



#8
 Diethyl Ether
 Concen: 54.785 ug/l
 RT: 3.534 min Scan# 281
 Delta R.T. 0.000 min
 Lab File: VY017197.D
 Acq: 02 Feb 2024 21:04

Tgt Ion: 74 Resp: 33545
 Ion Ratio Lower Upper
 74 100
 45 94.8 47.9 143.7









#9
1,1,2-Trichlorotrifluoroethane
Concen: 49.424 ug/l
RT: 3.906 min Scan# 341
Delta R.T. 0.006 min
Lab File: VY017197.D
Acq: 02 Feb 2024 21:04

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion:101 Resp: 58320
Ion Ratio Lower Upper
101 100
85 45.3 35.8 53.6
151 73.9 58.6 88.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 02/05/2024
Supervised By :Mahesh Dadoda 02/05/2024



#10
Methyl Iodide
Concen: 53.870 ug/l
RT: 4.095 min Scan# 373
Delta R.T. -0.006 min
Lab File: VY017197.D
Acq: 02 Feb 2024 21:04


Tgt Ion:142 Resp: 66828
Ion Ratio Lower Upper
142 100
127 41.8 35.1 52.7
141 14.4 11.8 17.6





#11
Tert butyl alcohol
Concen: 240.999 ug/l
RT: 4.960 min Scan# 515
Delta R.T. 0.000 min
Lab File: VY017197.D
Acq: 02 Feb 2024 21:04




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion: 59 Resp: 2072
Ion Ratio Lower Upper
59 100
57 9.8 7.4 11.0

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 02/05/2024
Supervised By :Mahesh Dadoda 02/05/2024




#12
1,1-Dichloroethene
Concen: 51.181 ug/l
RT: 3.875 min Scan# 337
Delta R.T. 0.000 min
Lab File: VY017197.D
Acq: 02 Feb 2024 21:04



Tgt Ion: 96 Resp: 56590
Ion Ratio Lower Upper
96 100
61 166.4 124.2 186.2
98 65.9 52.6 79.0





#13

Acrolein

Concen: 225.837 ug/l

RT: 3.735 min Scan# 314

Delta R.T. 0.001 min

Lab File: VY017197.D

Acq: 02 Feb 2024 21:04

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion: 56 Resp: 19164

Ion Ratio Lower Upper

56 100





55 69.9 55.4 83.0

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 02/05/2024

Supervised By :Mahesh Dadoda 02/05/2024



#14

Allyl chloride

Concen: 54.098 ug/l

RT: 4.485 min Scan# 437

Delta R.T. 0.000 min

Lab File: VY017197.D

Acq: 02 Feb 2024 21:04




Tgt Ion: 41 Resp: 90427


Ion Ratio Lower Upper

41 100

39 69.2 54.7 82.1


76 40.1 31.8 47.6





#15
Acrylonitrile
Concen: 268.358 ug/l
RT: 5.167 min Scan# 54
Delta R.T. 0.000 min
Lab File: VY017197.D
Acq: 02 Feb 2024 21:04




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 53 Resp: 7398
Ion Ratio Lower Upper
53 100
52 81.5 65.1 97.7
51 36.0 29.0 43.4




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 02/05/2024
Supervised By :Mahesh Dadoda 02/05/2024



#16
Acetone
Concen: 209.855 ug/l
RT: 3.954 min Scan# 350
Delta R.T. 0.006 min
Lab File: VY017197.D
Acq: 02 Feb 2024 21:04

Tgt Ion: 43 Resp: 62852
Ion Ratio Lower Upper
43 100
58 34.1 26.6 39.8









#17
Carbon Disulfide
Concen: 51.914 ug/l
RT: 4.198 min Scan# 391
Delta R.T. -0.006 min
Lab File: VY017197.D
Acq: 02 Feb 2024 21:04

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion: 76 Resp: 183229
Ion Ratio Lower Upper
76 100
78 9.1 7.4 11.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 02/05/2024
Supervised By :Mahesh Dadoda 02/05/2024



#18
Methyl Acetate
Concen: 62.116 ug/l
RT: 4.485 min Scan# 437
Delta R.T. 0.000 min
Lab File: VY017197.D
Acq: 02 Feb 2024 21:04


Tgt Ion: 43 Resp: 41502
Ion Ratio Lower Upper
43 100
74 24.5 19.8 29.6





#19
Methyl tert-butyl Ether
Concen: 54.039 ug/l
RT: 5.228 min Scan# 51
Delta R.T. 0.000 min
Lab File: VY017197.D
Acq: 02 Feb 2024 21:04




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 73 Resp: 14390
Ion Ratio Lower Upper
73 100
57 21.6 17.7 26.5




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 02/05/2024
Supervised By :Mahesh Dadoda 02/05/2024



#20
Methylene Chloride
Concen: 57.490 ug/l
RT: 4.722 min Scan# 476
Delta R.T. -0.006 min
Lab File: VY017197.D
Acq: 02 Feb 2024 21:04


Tgt Ion: 84 Resp: 69877
Ion Ratio Lower Upper
84 100
49 120.1 95.2 142.8
51 37.0 29.9 44.9
86 64.5 52.2 78.4





#21
trans-1,2-Dichloroethene
Concen: 54.720 ug/l
RT: 5.228 min Scan# 511
Delta R.T. 0.000 min
Lab File: VY017197.D
Acq: 02 Feb 2024 21:04




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 96 Resp: 68829
Ion Ratio Lower Upper
96 100
61 130.0 104.6 156.8
98 63.2 50.9 76.3




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 02/05/2024
Supervised By :Mahesh Dadoda 02/05/2024



#22
Diisopropyl ether
Concen: 55.799 ug/l
RT: 6.125 min Scan# 706
Delta R.T. 0.000 min
Lab File: VY017197.D
Acq: 02 Feb 2024 21:04


Tgt Ion: 45 Resp: 204825
Ion Ratio Lower Upper
45 100
43 55.5 46.8 70.2
87 27.4 22.2 33.2
59 11.9 9.9 14.9





#23
 Vinyl Acetate
 Concen: 262.954 ug/l
 RT: 6.070 min Scan# 697
 Delta R.T. 0.000 min
 Lab File: VY017197.D
 Acq: 02 Feb 2024 21:04




Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050




Tgt Ion: 43 Resp: 58485
 Ion Ratio Lower Upper
 43 100
 86 10.7 8.6 13.0

Manual Integrations
 APPROVED



Reviewed By :Semsettin Yesilyurt 02/05/2024
 Supervised By :Mahesh Dadoda 02/05/2024



#24
 1,1-Dichloroethane
 Concen: 54.215 ug/l
 RT: 6.027 min Scan# 690
 Delta R.T. 0.000 min
 Lab File: VY017197.D
 Acq: 02 Feb 2024 21:04



Tgt Ion: 63 Resp: 121913
 Ion Ratio Lower Upper
 63 100
 98 7.8 3.8 11.4
 100 5.0 2.1 6.2









#25
2-Butanone
Concen: 240.193 ug/l
RT: 6.990 min Scan# 848
Delta R.T. 0.000 min
Lab File: VY017197.D
Acq: 02 Feb 2024 21:04

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion: 43 Resp: 9563
Ion Ratio Lower Upper
43 100
72 28.5 21.4 32.2




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 02/05/2024
Supervised By :Mahesh Dadoda 02/05/2024



#26
2,2-Dichloropropane
Concen: 50.087 ug/l
RT: 6.990 min Scan# 848
Delta R.T. 0.000 min
Lab File: VY017197.D
Acq: 02 Feb 2024 21:04


Tgt Ion: 77 Resp: 93683
Ion Ratio Lower Upper
77 100
97 23.0 11.7 35.0





#27
 cis-1,2-Dichloroethene
 Concen: 54.622 ug/l
 RT: 6.990 min Scan# 849
 Delta R.T. -0.006 min
 Lab File: VY017197.D
 Acq: 02 Feb 2024 21:04




Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050



Tgt Ion: 96 Resp: 7684
 Ion Ratio Lower Upper
 96 100
 61 138.4 0.0 278.8
 98 65.1 0.0 127.0




Manual Integrations
 APPROVED


Reviewed By :Semsettin Yesilyurt 02/05/2024
 Supervised By :Mahesh Dadoda 02/05/2024



#28
 Bromochloromethane
 Concen: 55.917 ug/l
 RT: 7.344 min Scan# 906
 Delta R.T. 0.000 min
 Lab File: VY017197.D
 Acq: 02 Feb 2024 21:04


Tgt Ion: 49 Resp: 48145
 Ion Ratio Lower Upper
 49 100
 129 1.9 0.0 3.8
 130 75.5 59.8 89.8





#29
Tetrahydrofuran
Concen: 267.302 ug/l
RT: 7.356 min Scan# 908
Delta R.T. 0.000 min
Lab File: VY017197.D
Acq: 02 Feb 2024 21:04




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion: 42 Resp: 60120
Ion Ratio Lower Upper
42 100
72 46.1 36.6 54.8
71 42.1 33.8 50.6

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 02/05/2024
Supervised By :Mahesh Dadoda 02/05/2024




#30
Chloroform
Concen: 54.748 ug/l
RT: 7.515 min Scan# 934
Delta R.T. 0.000 min
Lab File: VY017197.D
Acq: 02 Feb 2024 21:04




Tgt Ion: 83 Resp: 123845
Ion Ratio Lower Upper
83 100
85 65.4 51.2 76.8





#31
Cyclohexane
Concen: 47.142 ug/l
RT: 7.795 min Scan# 980
Delta R.T. 0.000 min
Lab File: VY017197.D
Acq: 02 Feb 2024 21:04




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion: 56 Resp: 96820
Ion Ratio Lower Upper
56 100
69 35.7 27.0 40.6
84 89.3 71.8 107.6

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 02/05/2024
Supervised By :Mahesh Dadoda 02/05/2024




#32
1,1,1-Trichloroethane
Concen: 52.749 ug/l
RT: 7.710 min Scan# 966
Delta R.T. 0.000 min
Lab File: VY017197.D
Acq: 02 Feb 2024 21:04



Tgt Ion: 97 Resp: 102638
Ion Ratio Lower Upper
97 100
99 65.2 51.2 76.8
61 44.5 35.0 52.6









#33
1,2-Dichloroethane-d4
Concen: 52.018 ug/l
RT: 8.149 min Scan# 1038
Delta R.T. 0.000 min
Lab File: VY017197.D
Acq: 02 Feb 2024 21:04

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion: 65 Resp: 5300
Ion Ratio Lower Upper
65 100
67 53.5 0.0 109.0




Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/05/2024
Supervised By :Mahesh Dadoda 02/05/2024



#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 8.697 min Scan# 1128
Delta R.T. 0.000 min
Lab File: VY017197.D
Acq: 02 Feb 2024 21:04

Tgt Ion:114 Resp: 198926
Ion Ratio Lower Upper
114 100
63 19.7 0.0 37.4
88 14.9 0.0 29.0









#35
Dibromofluoromethane
Concen: 51.263 ug/l
RT: 7.722 min Scan# 969
Delta R.T. -0.006 min
Lab File: VY017197.D
Acq: 02 Feb 2024 21:04

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion:113 Resp: 55060
Ion Ratio Lower Upper
113 100
111 104.8 83.8 125.6
192 18.7 14.5 21.7




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 02/05/2024
Supervised By :Mahesh Dadoda 02/05/2024



#36
1,1-Dichloropropene
Concen: 51.120 ug/l
RT: 7.923 min Scan# 1001
Delta R.T. 0.000 min
Lab File: VY017197.D
Acq: 02 Feb 2024 21:04


Tgt Ion: 75 Resp: 94710
Ion Ratio Lower Upper
75 100
110 35.0 17.4 52.3
77 30.3 24.7 37.1





#37
Ethyl Acetate
Concen: 48.858 ug/l
RT: 7.082 min Scan# 863
Delta R.T. 0.000 min
Lab File: VY017197.D
Acq: 02 Feb 2024 21:04




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 43 Resp: 41984
Ion Ratio Lower Upper
43 100
61 14.7 11.0 16.6
70 9.9 8.1 12.1




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 02/05/2024
Supervised By :Mahesh Dadoda 02/05/2024



#38
Carbon Tetrachloride
Concen: 50.411 ug/l
RT: 7.911 min Scan# 999
Delta R.T. 0.000 min
Lab File: VY017197.D
Acq: 02 Feb 2024 21:04


Tgt Ion:117 Resp: 89424
Ion Ratio Lower Upper
117 100
119 95.4 76.0 114.0
121 30.7 24.2 36.2





#39
Methylcyclohexane
Concen: 48.905 ug/l
RT: 9.191 min Scan# 1209
Delta R.T. 0.000 min
Lab File: VY017197.D
Acq: 02 Feb 2024 21:04




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 83 Resp: 105853
Ion Ratio Lower Upper
83 100
55 73.6 59.7 89.5
98 49.5 38.0 57.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 02/05/2024
Supervised By :Mahesh Dadoda 02/05/2024



#40
Benzene
Concen: 53.216 ug/l
RT: 8.167 min Scan# 1041
Delta R.T. 0.000 min
Lab File: VY017197.D
Acq: 02 Feb 2024 21:04


Tgt Ion: 78 Resp: 281553
Ion Ratio Lower Upper
78 100
77 24.1 18.8 28.2





#41
Methacrylonitrile
Concen: 60.172 ug/l
RT: 7.320 min Scan# 902
Delta R.T. 0.000 min
Lab File: VY017197.D
Acq: 02 Feb 2024 21:04




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 41 Resp: 24498
Ion Ratio Lower Upper
41 100
39 51.1 41.9 62.9
67 74.6 59.8 89.6
52 29.0 24.8 37.2




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 02/05/2024
Supervised By :Mahesh Dadoda 02/05/2024



#42
1,2-Dichloroethane
Concen: 51.881 ug/l
RT: 8.246 min Scan# 1054
Delta R.T. 0.000 min
Lab File: VY017197.D
Acq: 02 Feb 2024 21:04


Tgt Ion: 62 Resp: 72676
Ion Ratio Lower Upper
62 100
98 10.3 0.0 20.2





#43
Isopropyl Acetate
Concen: 51.794 ug/l
RT: 8.283 min Scan# 1060
Delta R.T. 0.006 min
Lab File: VY017197.D
Acq: 02 Feb 2024 21:04




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 43 Resp: 78780
Ion Ratio Lower Upper
43 100
61 30.7 24.8 37.2
87 13.3 10.7 16.1




Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/05/2024
Supervised By :Mahesh Dadoda 02/05/2024



#44
Trichloroethene
Concen: 53.561 ug/l
RT: 8.947 min Scan# 1169
Delta R.T. 0.000 min
Lab File: VY017197.D
Acq: 02 Feb 2024 21:04

Tgt Ion:130 Resp: 71934
Ion Ratio Lower Upper
130 100
95 96.3 0.0 203.8





#45

1,2-Dichloropropane

Concen: 52.596 ug/l

RT: 9.222 min Scan# 1214

Delta R.T. -0.006 min

Lab File: VY017197.D

Acq: 02 Feb 2024 21:04

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion: 63 Resp: 6835

Ion Ratio Lower Upper

63 100





65 30.2 24.6 36.8

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 02/05/2024

Supervised By :Mahesh Dadoda 02/05/2024



#46

Dibromomethane

Concen: 52.445 ug/l

RT: 9.313 min Scan# 1229

Delta R.T. 0.000 min

Lab File: VY017197.D

Acq: 02 Feb 2024 21:04




Tgt Ion: 93 Resp: 36816


Ion Ratio Lower Upper

93 100

95 84.1 66.6 99.8


174 92.3 73.8 110.6





#47
Bromodichloromethane
Concen: 53.098 ug/l
RT: 9.502 min Scan# 1260
Delta R.T. 0.000 min
Lab File: VY017197.D
Acq: 02 Feb 2024 21:04




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 83 Resp: 93285
Ion Ratio Lower Upper
83 100
85 63.0 53.2 79.8
127 8.3 6.4 9.6




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 02/05/2024
Supervised By :Mahesh Dadoda 02/05/2024



#48
Methyl methacrylate
Concen: 51.725 ug/l
RT: 9.301 min Scan# 1227
Delta R.T. 0.000 min
Lab File: VY017197.D
Acq: 02 Feb 2024 21:04

Tgt Ion: 41 Resp: 36906
Ion Ratio Lower Upper
41 100
69 93.0 72.1 108.1
39 57.6 42.8 64.2





#49

1,4-Dioxane

Concen: 1083.912 ug/l

RT: 9.307 min Scan# 11

Delta R.T. 0.000 min

Lab File: VY017197.D

Acq: 02 Feb 2024 21:04

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion: 88 Resp: 828

Ion Ratio Lower Upper

88 100

43 33.2 28.2 42.4





58 62.8 52.2 78.2

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 02/05/2024

Supervised By :Mahesh Dadoda 02/05/2024



#50

Toluene-d8

Concen: 52.997 ug/l

RT: 10.185 min Scan# 1372

Delta R.T. 0.000 min

Lab File: VY017197.D




Acq: 02 Feb 2024 21:04


Tgt Ion: 98 Resp: 207251

Ion Ratio Lower Upper

98 100

100 65.5 52.3 78.5









#51
4-Methyl-2-Pentanone
Concen: 265.514 ug/l
RT: 10.081 min Scan# 1355
Delta R.T. 0.006 min
Lab File: VY017197.D
Acq: 02 Feb 2024 21:04

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion: 43 Resp: 216502
Ion Ratio Lower Upper
43 100
58 40.6 32.2 48.2




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 02/05/2024
Supervised By :Mahesh Dadoda 02/05/2024



#52
Toluene
Concen: 54.017 ug/l
RT: 10.252 min Scan# 1383
Delta R.T. 0.000 min
Lab File: VY017197.D
Acq: 02 Feb 2024 21:04


Tgt Ion: 92 Resp: 172507
Ion Ratio Lower Upper
92 100
91 173.4 137.6 206.4





#53
t-1,3-Dichloropropene
Concen: 52.666 ug/l
RT: 10.471 min Scan# 1419
Delta R.T. 0.000 min
Lab File: VY017197.D
Acq: 02 Feb 2024 21:04




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion: 75 Resp: 86909
Ion Ratio Lower Upper
75 100
77 31.8 24.6 36.8

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 02/05/2024
Supervised By :Mahesh Dadoda 02/05/2024




#54
cis-1,3-Dichloropropene
Concen: 53.006 ug/l
RT: 9.935 min Scan# 1331
Delta R.T. 0.000 min
Lab File: VY017197.D
Acq: 02 Feb 2024 21:04




Tgt Ion: 75 Resp: 104771
Ion Ratio Lower Upper
75 100
77 31.2 24.6 36.8
39 42.5 33.6 50.4





#55
1,1,2-Trichloroethane
Concen: 53.460 ug/l
RT: 10.654 min Scan# 1449
Delta R.T. 0.000 min
Lab File: VY017197.D
Acq: 02 Feb 2024 21:04




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion:	97	Resp:	51060
Ion	Ratio	Lower	Upper
97	100		
83	87.4	69.4	104.2
85	56.8	43.9	65.9
99	62.2	48.6	73.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 02/05/2024
Supervised By :Mahesh Dadoda 02/05/2024



#56
Ethyl methacrylate
Concen: 49.971 ug/l
RT: 10.520 min Scan# 1427
Delta R.T. 0.000 min
Lab File: VY017197.D
Acq: 02 Feb 2024 21:04


Tgt Ion:	69	Resp:	64213
Ion	Ratio	Lower	Upper
69	100		
41	61.7	49.2	73.8
39	32.5	26.8	40.2





#57
1,3-Dichloropropane
Concen: 53.795 ug/l
RT: 10.801 min Scan# 1473
Delta R.T. 0.000 min
Lab File: VY017197.D
Acq: 02 Feb 2024 21:04




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 76 Resp: 88372
Ion Ratio Lower Upper
76 100
78 31.8 26.1 39.1




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 02/05/2024
Supervised By :Mahesh Dadoda 02/05/2024



#58
2-Chloroethyl Vinyl ether
Concen: 286.159 ug/l
RT: 9.789 min Scan# 1307
Delta R.T. 0.000 min
Lab File: VY017197.D
Acq: 02 Feb 2024 21:04


Tgt Ion: 63 Resp: 170269
Ion Ratio Lower Upper
63 100
106 27.7 22.4 33.6





#59
2-Hexanone
Concen: 249.464 ug/l
RT: 10.837 min Scan# 1479
Delta R.T. 0.000 min
Lab File: VY017197.D
Acq: 02 Feb 2024 21:04




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 43 Resp: 147090
Ion Ratio Lower Upper
43 100
58 56.3 28.1 84.3




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 02/05/2024
Supervised By :Mahesh Dadoda 02/05/2024



#60
Dibromochloromethane
Concen: 52.898 ug/l
RT: 10.996 min Scan# 1505
Delta R.T. 0.006 min
Lab File: VY017197.D
Acq: 02 Feb 2024 21:04


Tgt Ion:129 Resp: 59271
Ion Ratio Lower Upper
129 100
127 78.6 39.2 117.6





#61
1,2-Dibromoethane
Concen: 52.917 ug/l
RT: 11.099 min Scan# 1522
Delta R.T. 0.000 min
Lab File: VY017197.D
Acq: 02 Feb 2024 21:04




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion:107 Resp: 47005
Ion Ratio Lower Upper
107 100
109 93.5 75.4 113.2

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 02/05/2024
Supervised By :Mahesh Dadoda 02/05/2024




#62
4-Bromofluorobenzene
Concen: 52.305 ug/l
RT: 12.489 min Scan# 1750
Delta R.T. 0.000 min
Lab File: VY017197.D
Acq: 02 Feb 2024 21:04




Tgt Ion: 95 Resp: 67005
Ion Ratio Lower Upper
95 100
174 81.5 0.0 160.0
176 77.6 0.0 151.8





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.502 min Scan# 117
Delta R.T. 0.000 min
Lab File: VY017197.D
Acq: 02 Feb 2024 21:04




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion:117 Resp: 173654
Ion Ratio Lower Upper
117 100
82 53.6 43.8 65.8
119 33.3 26.5 39.7




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 02/05/2024
Supervised By :Mahesh Dadoda 02/05/2024



#64
Tetrachloroethene
Concen: 53.971 ug/l
RT: 10.727 min Scan# 1461
Delta R.T. 0.000 min
Lab File: VY017197.D
Acq: 02 Feb 2024 21:04

Tgt Ion:164 Resp: 60868
Ion Ratio Lower Upper
164 100
166 130.3 98.8 148.2
129 98.4 78.9 118.3
131 93.0 72.5 108.7





#65

Chlorobenzene

Concen: 52.819 ug/l

RT: 11.526 min Scan# 1159

Delta R.T. 0.000 min

Lab File: VY017197.D

Acq: 02 Feb 2024 21:04

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion:112 Resp: 176939

Ion Ratio Lower Upper

112 100





114 33.4 26.1 39.1

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 02/05/2024

Supervised By :Mahesh Dadoda 02/05/2024



#66

1,1,1,2-Tetrachloroethane

Concen: 54.055 ug/l

RT: 11.599 min Scan# 1604

Delta R.T. 0.000 min

Lab File: VY017197.D

Acq: 02 Feb 2024 21:04




Tgt Ion:131 Resp: 65807


Ion Ratio Lower Upper

131 100

133 95.6 47.0 141.0


119 67.0 33.3 99.9





#67
Ethyl Benzene
Concen: 53.892 ug/l
RT: 11.599 min Scan# 1604
Delta R.T. 0.000 min
Lab File: VY017197.D
Acq: 02 Feb 2024 21:04




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion: 91 Resp: 32627
Ion Ratio Lower Upper
91 100
106 31.1 25.0 37.4

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 02/05/2024
Supervised By :Mahesh Dadoda 02/05/2024




#68
m/p-Xylenes
Concen: 108.951 ug/l
RT: 11.709 min Scan# 1622
Delta R.T. 0.000 min
Lab File: VY017197.D
Acq: 02 Feb 2024 21:04



Tgt Ion:106 Resp: 250324
Ion Ratio Lower Upper
106 100
91 201.0 160.7 241.1









#69
o-Xylene
Concen: 55.203 ug/l
RT: 12.038 min Scan# 1676
Delta R.T. 0.000 min
Lab File: VY017197.D
Acq: 02 Feb 2024 21:04

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion:106 Resp: 116800
Ion Ratio Lower Upper
106 100
91 212.6 106.5 319.5




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 02/05/2024
Supervised By :Mahesh Dadoda 02/05/2024



#70
Styrene
Concen: 56.394 ug/l
RT: 12.050 min Scan# 1678
Delta R.T. 0.000 min
Lab File: VY017197.D
Acq: 02 Feb 2024 21:04

Tgt Ion:104 Resp: 196595
Ion Ratio Lower Upper
104 100
78 47.7 39.2 58.8
103 54.7 43.8 65.6





#71

Bromoform

Concen: 53.269 ug/l

RT: 12.215 min Scan# 11

Delta R.T. 0.000 min

Lab File: VY017197.D

Acq: 02 Feb 2024 21:04

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion:173 Resp: 33962

Ion Ratio Lower Upper

173 100

175 48.1 23.5 70.6





254 0.0 0.0 0.0

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 02/05/2024

Supervised By :Mahesh Dadoda 02/05/2024



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.434 min Scan# 1905

Delta R.T. 0.000 min

Lab File: VY017197.D

Acq: 02 Feb 2024 21:04




Tgt Ion:152 Resp: 79382


Ion Ratio Lower Upper

152 100

115 58.3 30.0 90.0

150 173.6 0.0 344.6





#73

Isopropylbenzene

Concen: 52.251 ug/l

RT: 12.337 min Scan# 1725

Delta R.T. 0.000 min

Lab File: VY017197.D

Acq: 02 Feb 2024 21:04

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion:105 Resp: 304710

Ion Ratio Lower Upper

105 100





120 26.2 13.0 39.0

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 02/05/2024

Supervised By :Mahesh Dadoda 02/05/2024



#74

N-ethyl acetate

Concen: 51.625 ug/l

RT: 12.154 min Scan# 1695

Delta R.T. 0.000 min

Lab File: VY017197.D

Acq: 02 Feb 2024 21:04

Tgt Ion: 43 Resp: 67647




Ion Ratio Lower Upper


43 100

70 44.2 36.6 55.0

55 27.6 22.2 33.4

61 24.8 20.2 30.4









#75
1,1,2,2-Tetrachloroethane
Concen: 49.839 ug/l
RT: 12.593 min Scan# 1766
Delta R.T. 0.006 min
Lab File: VY017197.D
Acq: 02 Feb 2024 21:04

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion: 83 Resp: 5716
Ion Ratio Lower Upper
83 100
131 9.4 4.9 14.6
85 64.6 32.6 98.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 02/05/2024
Supervised By :Mahesh Dadoda 02/05/2024



#76
1,2,3-Trichloropropane
Concen: 51.459 ug/l m
RT: 12.642 min Scan# 1775
Delta R.T. 0.000 min
Lab File: VY017197.D
Acq: 02 Feb 2024 21:04

Tgt Ion: 75 Resp: 37703
Ion Ratio Lower Upper
75 100
77 0.0 0.0 0.0





#77

Bromobenzene

Concen: 52.607 ug/l

RT: 12.617 min Scan# 1771

Delta R.T. 0.000 min

Lab File: VY017197.D

Acq: 02 Feb 2024 21:04

Instrument :

MSVOA_Y

ClientSampleId :





VSTDCCC050

Tgt Ion:	156	Resp:	68050
Ion Ratio	Lower	Upper	
156	100		
77	186.2	97.3	291.8
158	95.0	48.4	145.0

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 02/05/2024
Supervised By :Mahesh Dadoda 02/05/2024



#78

n-propylbenzene

Concen: 52.379 ug/l




RT: 12.678 min Scan# 1781


Delta R.T. 0.000 min

Lab File: VY017197.D

Acq: 02 Feb 2024 21:04


Tgt Ion:	91	Resp:	375160
Ion Ratio	Lower	Upper	
91	100		
120	22.2	11.0	33.0





#79
2-Chlorotoluene
Concen: 52.683 ug/l
RT: 12.764 min Scan# 1795
Delta R.T. 0.000 min
Lab File: VY017197.D
Acq: 02 Feb 2024 21:04




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion: 91 Resp: 207449
Ion Ratio Lower Upper
91 100
126 34.6 17.0 51.0

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 02/05/2024
Supervised By :Mahesh Dadoda 02/05/2024




#80
1,3,5-Trimethylbenzene
Concen: 53.343 ug/l
RT: 12.825 min Scan# 1805
Delta R.T. 0.006 min
Lab File: VY017197.D
Acq: 02 Feb 2024 21:04



Tgt Ion:105 Resp: 247405
Ion Ratio Lower Upper
105 100
120 49.5 24.3 72.8









#81
trans-1,4-Dichloro-2-butene
Concen: 48.787 ug/l
RT: 12.386 min Scan# 1733
Delta R.T. 0.000 min
Lab File: VY017197.D
Acq: 02 Feb 2024 21:04

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion: 75 Resp: 1720
Ion Ratio Lower Upper
75 100
53 92.4 74.0 111.0
89 43.5 36.5 54.7




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 02/05/2024
Supervised By :Mahesh Dadoda 02/05/2024



#82
4-Chlorotoluene
Concen: 52.176 ug/l
RT: 12.867 min Scan# 1812
Delta R.T. 0.006 min
Lab File: VY017197.D
Acq: 02 Feb 2024 21:04


Tgt Ion: 91 Resp: 211449
Ion Ratio Lower Upper
91 100
126 34.8 17.1 51.2





#83
tert-Butylbenzene
Concen: 54.226 ug/l
RT: 13.087 min Scan# 1847
Delta R.T. 0.006 min
Lab File: VY017197.D
Acq: 02 Feb 2024 21:04




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion:119 Resp: 21558
Ion Ratio Lower Upper
119 100
91 62.3 31.5 94.5
134 25.9 13.1 39.3




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 02/05/2024
Supervised By :Mahesh Dadoda 02/05/2024



#84
1,2,4-Trimethylbenzene
Concen: 53.576 ug/l
RT: 13.129 min Scan# 1855
Delta R.T. 0.000 min
Lab File: VY017197.D
Acq: 02 Feb 2024 21:04

Tgt Ion:105 Resp: 245271
Ion Ratio Lower Upper
105 100
120 44.9 22.1 66.3





#85

sec-Butylbenzene

Concen: 51.629 ug/l

RT: 13.264 min Scan# 1877

Delta R.T. 0.000 min

Lab File: VY017197.D

Acq: 02 Feb 2024 21:04

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion:105 Resp: 32080

Ion Ratio Lower Upper

105 100





134 19.8 9.7 28.9

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 02/05/2024

Supervised By :Mahesh Dadoda 02/05/2024



#86

p-Isopropyltoluene

Concen: 52.511 ug/l

RT: 13.379 min Scan# 1896

Delta R.T. 0.000 min

Lab File: VY017197.D

Acq: 02 Feb 2024 21:04




Tgt Ion:119 Resp: 260738


Ion Ratio Lower Upper

119 100

134 26.4 13.1 39.1


91 22.9 11.7 35.0





#87
1,3-Dichlorobenzene
Concen: 52.226 ug/l
RT: 13.373 min Scan# 1895
Delta R.T. 0.000 min
Lab File: VY017197.D
Acq: 02 Feb 2024 21:04




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion:146 Resp: 134770
Ion Ratio Lower Upper
146 100
111 39.2 20.2 60.5
148 63.7 31.9 95.9

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 02/05/2024
Supervised By :Mahesh Dadoda 02/05/2024




#88
1,4-Dichlorobenzene
Concen: 52.890 ug/l
RT: 13.459 min Scan# 1909
Delta R.T. 0.006 min
Lab File: VY017197.D
Acq: 02 Feb 2024 21:04




Tgt Ion:146 Resp: 132598
Ion Ratio Lower Upper
146 100
111 38.5 19.3 57.9
148 63.2 31.6 94.7





#89
n-Butylbenzene
Concen: 52.468 ug/l
RT: 13.703 min Scan# 1949
Delta R.T. 0.000 min
Lab File: VY017197.D
Acq: 02 Feb 2024 21:04




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 91 Resp: 24770
Ion Ratio Lower Upper
91 100
92 52.0 26.4 79.2
134 24.5 12.3 36.8




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 02/05/2024
Supervised By :Mahesh Dadoda 02/05/2024



#90
Hexachloroethane
Concen: 51.431 ug/l
RT: 13.971 min Scan# 1993
Delta R.T. 0.000 min
Lab File: VY017197.D
Acq: 02 Feb 2024 21:04


Tgt Ion:117 Resp: 52426
Ion Ratio Lower Upper
117 100
201 68.0 33.4 100.1





#91
1,2-Dichlorobenzene
Concen: 53.187 ug/l
RT: 13.751 min Scan# 1957
Delta R.T. 0.006 min
Lab File: VY017197.D
Acq: 02 Feb 2024 21:04




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion:146 Resp: 11684
Ion Ratio Lower Upper
146 100
111 41.6 20.5 61.6
148 64.1 31.4 94.3

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 02/05/2024
Supervised By :Mahesh Dadoda 02/05/2024




#92
1,2-Dibromo-3-Chloropropane
Concen: 49.825 ug/l
RT: 14.367 min Scan# 2058
Delta R.T. 0.006 min
Lab File: VY017197.D
Acq: 02 Feb 2024 21:04



Tgt Ion: 75 Resp: 7444
Ion Ratio Lower Upper
75 100
155 88.9 42.1 126.3
157 109.7 55.8 167.4





#93

1,2,4-Trichlorobenzene

Concen: 56.760 ug/l

RT: 15.019 min Scan# 2165

Delta R.T. 0.006 min

Lab File: VY017197.D

Acq: 02 Feb 2024 21:04

Instrument :

MSVOA_Y

ClientSampleId :





VSTDCCC050

Tgt Ion:180	Resp: 6193
Ion Ratio	Lower Upper
180 100	
182 95.0	47.9 143.7
145 30.4	15.9 47.7

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 02/05/2024
Supervised By :Mahesh Dadoda 02/05/2024



#94

Hexachlorobutadiene

Concen: 52.046 ug/l




RT: 15.123 min Scan# 2182


Delta R.T. 0.006 min

Lab File: VY017197.D

Acq: 02 Feb 2024 21:04


Tgt Ion:225	Resp: 33448
Ion Ratio	Lower Upper
225 100	
223 62.6	31.4 94.3
227 65.7	32.3 96.8





#95
Naphthalene
Concen: 52.674 ug/l
RT: 15.245 min Scan# 2202
Delta R.T. -0.000 min
Lab File: VY017197.D
Acq: 02 Feb 2024 21:04




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion:128 Resp: 11684
Ion Ratio Lower Upper
128 100
127 12.6 10.2 15.4
129 10.8 8.6 12.8

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/05/2024
Supervised By :Mahesh Dadoda 02/05/2024



#96
1,2,3-Trichlorobenzene
Concen: 56.947 ug/l
RT: 15.434 min Scan# 2233
Delta R.T. 0.000 min
Lab File: VY017197.D
Acq: 02 Feb 2024 21:04



Tgt Ion:180 Resp: 52849
Ion Ratio Lower Upper
180 100
182 95.7 46.3 138.8
145 32.3 16.1 48.2

