



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020224\
Data File : VY017174.D
Acq On : 02 Feb 2024 08:31
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020124S.M
Title : SW846 8260
Last Update : Fri Feb 02 00:56:04 2024



AutoFind: Scans 1750, 1751, 1752; Background Corrected with Scan 1742

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	20.8	14982	PASS
75	95	30	60	53.6	38608	PASS
95	95	100	100	100.0	72021	PASS
96	95	5	9	6.8	4870	PASS
173	174	0.00	2	1.3	695	PASS
174	95	50	100	76.0	54712	PASS
175	174	5	9	7.8	4243	PASS
176	174	95	101	95.4	52208	PASS
177	176	5	9	6.5	3375	PASS