


Data Path : Z:\voasrv\HPCHEM1\MSVOA Y\Data\vv020320\
 Data File : VY001462.D
 Acq On : 03 Feb 2020 10:03
 Operator : SY/MD
 Sample : BFB
 Misc : 5.00G/5ML/MSVOA Y/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_Y\METHODS\82Y012120S.M
 Title : SW846 8260
 Last Update : Thu Jan 23 01:21:47 2020



AutoFind: Scans 1762, 1763, 1764; Background Corrected with Scan 1753

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.2	28728	PASS
75	95	30	60	50.2	75016	PASS
95	95	100	100	100.0	149312	PASS
96	95	5	9	6.9	10346	PASS
173	174	0.00	2	0.8	844	PASS
174	95	50	100	72.7	108528	PASS
175	174	5	9	7.4	8020	PASS
176	174	95	101	95.8	103976	PASS
177	176	5	9	6.8	7024	PASS