

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020321\
 Data File : VY003964.D
 Acq On : 03 Feb 2021 17:05
 Operator : SY/MD
 Sample : M1292-17
 Misc : 6.52g/5mL/MSVOA_Y/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 2021-01-28-FD04

Quant Time: Feb 04 01:21:05 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y012821S.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 01 00:38:02 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	90908	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	153650	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	144361	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	73191	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.155	65	61479	61.50	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 123.00%	
35) Dibromofluoromethane	7.728	113	50174	55.85	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 111.70%	
50) Toluene-d8	10.185	98	186066	52.78	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 105.56%	
62) 4-Bromofluorobenzene	12.483	95	69790	54.31	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 108.62%	
Target Compounds						
16) Acetone	3.960	43	10179	41.48	ug/l	98
17) Carbon Disulfide	4.210	76	4160	1.54	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020321\
Data File : VY003964.D
Acq On : 03 Feb 2021 17:05
Operator : SY/MD
Sample : M1292-17
Misc : 6.52g/5mL/MSVOA_Y/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
2021-01-28-FD04

Quant Time: Feb 04 01:21:05 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y012821S.M
Quant Title : SW846 8260
QLast Update : Mon Feb 01 00:38:02 2021
Response via : Initial Calibration

