

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020321\
 Data File : VY003973.D
 Acq On : 03 Feb 2021 20:35
 Operator : SY/MD
 Sample : M1291-09
 Misc : 7.42g/5mL/MSVOA_Y/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 2021-01-27-FD03

Quant Time: Feb 04 01:23:33 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y012821S.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 01 00:38:02 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	82741	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	136459	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	132363	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	67531	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.155	65	55437	60.93	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 121.86%	
35) Dibromofluoromethane	7.728	113	44410	55.66	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 111.32%	
50) Toluene-d8	10.185	98	168118	53.69	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 107.38%	
62) 4-Bromofluorobenzene	12.483	95	63512	55.65	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 111.30%	
Target Compounds						Qvalue
16) Acetone	3.966	43	4286	19.19	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020321\
Data File : VY003973.D
Acq On : 03 Feb 2021 20:35
Operator : SY/MD
Sample : M1291-09
Misc : 7.42g/5mL/MSVOA_Y/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
2021-01-27-FD03

Quant Time: Feb 04 01:23:33 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y012821S.M
Quant Title : SW846 8260
QLast Update : Mon Feb 01 00:38:02 2021
Response via : Initial Calibration

