

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020321\
 Data File : VY003980.D
 Acq On : 03 Feb 2021 23:17
 Operator : SY/MD
 Sample : M1291-16
 Misc : 8.67g/5mL/MSVOA_Y/SOIL
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 OU421-004-0002

Quant Time: Feb 04 01:25:57 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y012821S.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 01 00:38:02 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	82399	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	134786	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	127795	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	63778	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.155	65	55072	60.78	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	121.56%
35) Dibromofluoromethane	7.728	113	43959	55.78	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	111.56%
50) Toluene-d8	10.185	98	164465	53.18	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	106.36%
62) 4-Bromofluorobenzene	12.483	95	61104	54.20	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	108.40%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020321\
Data File : VY003980.D
Acq On : 03 Feb 2021 23:17
Operator : SY/MD
Sample : M1291-16
Misc : 8.67g/5mL/MSVOA_Y/SOIL
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
OU421-004-0002

Quant Time: Feb 04 01:25:57 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y012821S.M
Quant Title : SW846 8260
QLast Update : Mon Feb 01 00:38:02 2021
Response via : Initial Calibration

