

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020323\
 Data File : VY012475.D
 Acq On : 03 Feb 2023 14:26
 Operator : KP/MD
 Sample : 01418-09
 Misc : 6.21g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 JPP-20-020223

Quant Time: Feb 04 00:09:26 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011623S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 04:31:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	169813	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	283701	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	261808	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	99371	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	105091	76.767	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	153.540%
35) Dibromofluoromethane	7.716	113	86525	57.599	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	115.200%
50) Toluene-d8	10.179	98	338846	66.535	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	133.060%
62) 4-Bromofluorobenzene	12.483	95	97891	46.226	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	92.460%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020323\
Data File : VY012475.D
Acq On : 03 Feb 2023 14:26
Operator : KP/MD
Sample : 01418-09
Misc : 6.21g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
JPP-20-020223

Quant Time: Feb 04 00:09:26 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011623S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 17 04:31:47 2023
Response via : Initial Calibration

