

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020323\
 Data File : VY012482.D
 Acq On : 03 Feb 2023 17:10
 Operator : KP/MD
 Sample : 01414-05
 Misc : 5.07g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SASP2-123-2(65-72)

Quant Time: Feb 04 00:12:13 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011623S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 04:31:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	172121	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	295428	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	280835	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	122124	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	103736	74.649	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	149.300%
35) Dibromofluoromethane	7.716	113	87446	55.901	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	111.800%
50) Toluene-d8	10.179	98	343422	64.630	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	129.260%
62) 4-Bromofluorobenzene	12.483	95	112544	51.036	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	102.080%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020323\
Data File : VY012482.D
Acq On : 03 Feb 2023 17:10
Operator : KP/MD
Sample : 01414-05
Misc : 5.07g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SASP2-123-2(65-72)

Quant Time: Feb 04 00:12:13 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011623S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 17 04:31:47 2023
Response via : Initial Calibration

