

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020323\
 Data File : VY012484.D
 Acq On : 03 Feb 2023 17:57
 Operator : KP/MD
 Sample : 01418-13
 Misc : 7.17g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 JPP-22-020223

Quant Time: Feb 04 00:13:32 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011623S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 04:31:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	166660	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	279065	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	256606	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	97697	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	101997	75.869	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	151.740%
35) Dibromofluoromethane	7.715	113	84471	57.166	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	114.340%
50) Toluene-d8	10.178	98	333267	66.526	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	133.060%
62) 4-Bromofluorobenzene	12.483	95	95395	45.796	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	91.600%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020323\
Data File : VY012484.D
Acq On : 03 Feb 2023 17:57
Operator : KP/MD
Sample : 01418-13
Misc : 7.17g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
JPP-22-020223

Quant Time: Feb 04 00:13:32 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011623S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 17 04:31:47 2023
Response via : Initial Calibration

