

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020325\
 Data File : VY021031.D
 Acq On : 03 Feb 2025 15:50
 Operator : SY/MD
 Sample : Q1239-10
 Misc : 5.06g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 357

Quant Time: Feb 04 00:56:31 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020325S.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 03 13:08:38 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	4138	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.622	114	6912	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	3829	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.359	152	420	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	2606	61.466	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	122.940%
35) Dibromofluoromethane	7.640	113	2495	56.325	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	112.640%
50) Toluene-d8	10.109	98	7177	42.558	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	85.120%
62) 4-Bromofluorobenzene	12.414	95	749	13.595	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	27.180%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020325\
Data File : VY021031.D
Acq On : 03 Feb 2025 15:50
Operator : SY/MD
Sample : Q1239-10
Misc : 5.06g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
357

Quant Time: Feb 04 00:56:31 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020325S.M
Quant Title : SW846 8260
QLast Update : Mon Feb 03 13:08:38 2025
Response via : Initial Calibration

