

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020325\
 Data File : VY021037.D
 Acq On : 03 Feb 2025 18:11
 Operator : SY/MD
 Sample : IBLK
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 IBLK

Quant Time: Feb 04 00:58:31 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020325S.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 03 13:08:38 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	0.000	168	0	0.000	ug/l	-7.71
34) 1,4-Difluorobenzene	8.622	114	27	50.000	ug/l	0.00
63) Chlorobenzene-d5	0.000	117	0	0.000	ug/l	-11.42
72) 1,4-Dichlorobenzene-d4	0.000	152	0	0.000	ug/l	-13.35
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	0.000	65	0	0.000	ug/l	
Spiked Amount	50.000	Range	50 - 163	Recovery	=	0.000%#
35) Dibromofluoromethane	0.000	113	0	0.000	ug/l	
Spiked Amount	50.000	Range	54 - 147	Recovery	=	0.000%#
50) Toluene-d8	10.109	98	57	86.527	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	173.060%#
62) 4-Bromofluorobenzene	0.000	95	0	0.000	ug/l	
Spiked Amount	50.000	Range	30 - 143	Recovery	=	0.000%#
Target Compounds						
					Qvalue	
51) 4-Methyl-2-Pentanone	9.841	43	19	169.122	ug/l #	35
59) 2-Hexanone	10.810	43	23	317.968	ug/l #	23

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020325\
Data File : VY021037.D
Acq On : 03 Feb 2025 18:11
Operator : SY/MD
Sample : IBLK
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
IBLK

Quant Time: Feb 04 00:58:31 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020325S.M
Quant Title : SW846 8260
QLast Update : Mon Feb 03 13:08:38 2025
Response via : Initial Calibration

