

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020422\
 Data File : VY007387.D
 Acq On : 04 Feb 2022 11:36
 Operator : SY/MD
 Sample : VY0204SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0204SBL01

Quant Time: Feb 06 23:32:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020222S.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 03 04:36:48 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	113609	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	208669	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	198984	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	85461	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	67244	51.458	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 102.920%	
35) Dibromofluoromethane	7.716	113	68038	50.777	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 101.560%	
50) Toluene-d8	10.179	98	259443	52.408	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 104.820%	
62) 4-Bromofluorobenzene	12.477	95	90644	53.871	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 107.740%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020422\
Data File : VY007387.D
Acq On : 04 Feb 2022 11:36
Operator : SY/MD
Sample : VY0204SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0204SBL01

Quant Time: Feb 06 23:32:31 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020222S.M
Quant Title : SW846 8260
QLast Update : Thu Feb 03 04:36:48 2022
Response via : Initial Calibration

