

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020422\
 Data File : VY007394.D
 Acq On : 04 Feb 2022 15:03
 Operator : SY/MD
 Sample : N1391-04
 Misc : 5.04g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TR-08-020322

Quant Time: Feb 06 23:35:40 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020222S.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 03 04:36:48 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	26226	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	48081	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	44206	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	18680	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	16185	53.653	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 107.300%	
35) Dibromofluoromethane	7.722	113	15327	49.643	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 99.280%	
50) Toluene-d8	10.179	98	59000	51.724	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 103.440%	
62) 4-Bromofluorobenzene	12.477	95	20307	52.378	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 104.760%	
Target Compounds						
						Qvalue
16) Acetone	3.948	43	3238	33.050	ug/l	95
52) Toluene	10.246	92	2844	3.155	ug/l	95
68) m/p-Xylenes	11.697	106	1599	2.314	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020422\
Data File : VY007394.D
Acq On : 04 Feb 2022 15:03
Operator : SY/MD
Sample : N1391-04
Misc : 5.04g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TR-08-020322

Quant Time: Feb 06 23:35:40 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020222S.M
Quant Title : SW846 8260
QLast Update : Thu Feb 03 04:36:48 2022
Response via : Initial Calibration

