

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020422\
 Data File : VY007397.D
 Acq On : 04 Feb 2022 16:13
 Operator : SY/MD
 Sample : N1387-05
 Misc : 5.07g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 26601

Quant Time: Feb 06 23:36:39 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020222S.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 03 04:36:48 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	0.000	168	0	0.000	ug/l	-7.79
34) 1,4-Difluorobenzene	0.000	114	0	0.000	ug/l	-8.69
63) Chlorobenzene-d5	0.000	117	0	0.000	ug/l	-11.49
72) 1,4-Dichlorobenzene-d4	0.000	152	0	0.000	ug/l	-13.42
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	0.000	65	0	0.000	ug/l	
Spiked Amount	50.000	Range	50 - 163	Recovery	=	0.000%#
35) Dibromofluoromethane	0.000	113	0	0.000	ug/l	
Spiked Amount	50.000	Range	54 - 147	Recovery	=	0.000%#
50) Toluene-d8	0.000	98	0	0.000	ug/l	
Spiked Amount	50.000	Range	49 - 140	Recovery	=	0.000%#
62) 4-Bromofluorobenzene	12.477	95	44	0.000	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	0.000%#

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY020422\
Data File : VY007397.D
Acq On : 04 Feb 2022 16:13
Operator : SY/MD
Sample : N1387-05
Misc : 5.07g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
26601

Quant Time: Feb 06 23:36:39 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y020222S.M
Quant Title : SW846 8260
QLast Update : Thu Feb 03 04:36:48 2022
Response via : Initial Calibration

